

Contract No:

This document was prepared in conjunction with work accomplished under Contract No. 89303321CEM000080 with the U.S. Department of Energy (DOE) Office of Environmental Management (EM).

Disclaimer:

This work was prepared under an agreement with and funded by the U.S. Government. Neither the U.S. Government or its employees, nor any of its contractors, subcontractors or their employees, makes any express or implied:

- 1) warranty or assumes any legal liability for the accuracy, completeness, or for the use or results of such use of any information, product, or process disclosed; or
- 2) representation that such use or results of such use would not infringe privately owned rights; or
- 3) endorsement or recommendation of any specifically identified commercial product, process, or service.

Any views and opinions of authors expressed in this work do not necessarily state or reflect those of the United States Government, or its contractors, or subcontractors.



**Savannah River
National Laboratory®**

A U.S. DEPARTMENT OF ENERGY NATIONAL LAB • SAVANNAH RIVER SITE • AIKEN, SC • USA

Performance Testing of Software Upgrade for SRNL CPC instrument at the Nuclear Material Laboratory, Office of Safeguards Analytical Services, International Atomic Energy Agency

Maria E. Morales-Arteaga, Principal Scientist, Chemical Process & Characterization

Andrew J. Kruzner, Software Engineer, Intelligence

Joseph V. Cordaro, Electrical Engineer, Global Security

Savannah River National Laboratory

May 2022

SRNL--L3100-2022-00021, Revision 0

SRNL.DOE.GOV

DISCLAIMER

This work was prepared under an agreement with and funded by the U.S. Government. Neither the U.S. Government or its employees, nor any of its contractors, subcontractors or their employees, makes any express or implied:

1. warranty or assumes any legal liability for the accuracy, completeness, or for the use or results of such use of any information, product, or process disclosed; or
2. representation that such use or results of such use would not infringe privately owned rights; or
3. endorsement or recommendation of any specifically identified commercial product, process, or service.

Any views and opinions of authors expressed in this work do not necessarily state or reflect those of the United States Government, or its contractors, or subcontractors.

Printed in the United States of America

**Prepared for
U.S. Department of Energy**

Keywords: *Software, Controlled Potential
Coulometry, performance testing*

Retention: *Varies*

**Performance Testing of Software
Upgrade for SRNL CPC instrument at the Nuclear
Material Laboratory, Office of Safeguards Analytical
Services, International Atomic Energy Agency**

Maria E. Morales-Arteaga

Andrew J. Kruzner

Joseph V. Cordaro

Savannah River National Laboratory

May 2022

Savannah River National Laboratory is operated by
Battelle Savannah River Alliance for the U.S. Department
of Energy under Contract No. 89303321CEM000080.



PREFACE OR ACKNOWLEDGEMENTS

Thanks to the IAEA and the US Support Program for Sponsoring the Software Upgrade.

EXECUTIVE SUMMARY

This report summarizes the performance testing for the Software upgrade for the Savannah River National Laboratory (SRNL) Controlled Potential Coulometry (CPC) instrument. This upgrade is specifically to meet the needs of the Nuclear Material Laboratory, Office of Safeguards Analytical Services, International Atomic Energy Agency at Seibersdorf, Austria. The instrument hardware was upgraded in 2010, but the software continued to operate on a Windows platform using a High Tech Basic (HT Basic) application. HT Basic usage is declining. The IAEA requested an upgrade to the existing coulometer software to Laboratory Virtual Instrument Engineering Workbench (LabVIEW), an object-oriented programming language in wide use and one the NML staff is familiar with. The software was developed by SRNL Software Engineer, with assistance from Electrical Engineer and Scientist. Once developed, the new LabVIEW software was validated, tested with Iron (surrogate for plutonium) and plutonium. Once testing was completed, the software was installed at the International Atomic Energy Agency Nuclear Material Laboratory (IAEA NML) coulometer, and SRNL provided hands on training to NML staff on using the new software.

TABLE OF CONTENTS

LIST OF TABLES	vii
LIST OF FIGURES	vii
LIST OF ABBREVIATIONS.....	viii
1.0 Introduction.....	1
2.0 Scope.....	1
3.0 Procedure	1
3.1 System Automatic Continuous Calibration.....	1
3.2 Iron Blank/Sample Measurements	1
3.2.1 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.....	1
3.2.2 Using Iron, demonstrate the running of a sample.	1
3.3 Plutonium Blank/Sample Measurements	1
3.3.1 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.....	2
3.3.2 Using Plutonium, demonstrate the running of a sample.	2
3.4 Creation of files for data archiving.....	2
3.5 Installation and operation at NML CPC.....	2
3.5.1 Perform continuous electrical calibrations	2
3.5.2 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.....	2
3.5.3 Using Plutonium, demonstrate the running of a sample.	2
4.0 Results and Discussion	2
4.1 System Automatic Continuous Calibration.....	2
4.2 Iron Blank/Sample Measurements	2
4.3 Plutonium Blank/Sample Measurements	3
4.4 Creation of files for data archiving.....	3
4.5 Installation and operation at NML CPC.....	4
5.0 Conclusions.....	4
6.0 Recommendations, Path Forward or Future Work	4
Appendix A . Calibration Report	A-1

LIST OF TABLES

Table 4-1. Summary of plutonium measurements.	3
Table 4-2. Main files generated by software.	3

LIST OF FIGURES

Figure 4-1. Example of output to Calibration text file.	2
Figure 4-2. Files created by new LabVIEW program on Data folder.	4

LIST OF ABBREVIATIONS

SRNL	Savannah River National Laboratory
CPC	Controlled Potential Coulometry
IAEA	International Atomic Energy Agency
NML	Nuclear Material Laboratory
LabVIEW	Laboratory Virtual Instrument Engineering Workbench
High Tech Basic	HT Basic

1.0 Introduction

SRNL has been fabricating CPC instruments for over 30 years. The instrument hardware was upgraded in 2010, but the software continued to operate on a Windows platform using a High Tech Basic (HT Basic) application. HT Basic usage is declining. The IAEA requested an upgrade to the existing coulometer software to Laboratory Virtual Instrument Engineering Workbench (LabVIEW), an object-oriented programming language in wide use and one the NML staff is familiar with. The software was developed by SRNL Software Engineer, with assistance from Electrical Engineer and chemist. Once developed, the new LabVIEW software was validated, testing with Iron (surrogate for plutonium) and plutonium. Once testing was completed, the software was installed on the International Atomic Energy Agency Nuclear Material Laboratory (IAEA NML) coulometer, and SRNL provided hands on training to NML staff using the new software.

2.0 Scope

The performance testing of the LabVIEW software for SRNL CPC will demonstrate acceptable operation of the coulometer by performing:

- System Automatic Continuous Calibration
- Iron Blank/Sample Measurements
- Plutonium Blank/Sample Measurements
- Creation of files for data archiving
- Installation and operation at NML CPC

The expected outcome is that the new software will produce results that will be equivalent to the results obtained using the HT Basic Program.

3.0 Procedure

3.1 System Automatic Continuous Calibration

Using the automatic calibration routine in the software, run at least three series of 10 calibration factors. The calibration file should appear similar to the calibration file for HT Basic. The system must be on for a minimum of 1 hour before starting calibration runs.

3.2 Iron Blank/Sample Measurements

Demonstrate Blank/Sample Measurement using Sulfuric acid solutions and Iron using the SRNL Cell. This test is not intended to test the performance of the cell assembly, acids, and iron. The purpose of this test is to determine if the software is functioning as expected. Note that system parameters may require changing to complete testing. This testing is only performed at SRNL.

3.2.1 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.

3.2.2 Using Iron, demonstrate the running of a sample.

3.3 Plutonium Blank/Sample Measurements

Demonstrate Blank/Sample Measurement using Sulfuric acid solutions and plutonium using the SRNL Cell. The purpose of this test is to determine if the software is functioning as expected and producing expected results, comparable to the HT Basic Software. Note that system parameters may require changing to complete testing.

3.3.1 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.

3.3.2 Using Plutonium, demonstrate the running of a sample.

3.4 Creation of files for data archiving

Demonstrate that the LabVIEW program creates the required txt and CXL files to archive data.

3.5 Installation and operation at NML CPC

Install the LabVIEW program on IAEA NML computer. Demonstrate Blank/Sample Measurement using Sulfuric acid solutions and plutonium using the IAEA NML Cell. The purpose of this test is to determine if the software is functioning as expected and producing expected results, comparable to the HT Basic Software. Note that system parameters may require changing to complete testing.

3.5.1 Perform continuous electrical calibrations

3.5.2 Using a sulfuric or nitric acid with an Argon purge, demonstrate the running of a blank.

3.5.3 Using Plutonium, demonstrate the running of a sample.

4.0 Results and Discussion

4.1 System Automatic Continuous Calibration

Using the automatic calibration routine in the software, the instrument was left running automatic continuous calibration. The calibration file produced was similar to the calibration file for HT Basic. See Figure 4-1.

Routine Electrical Calibration Process -- Data are saved as tab-separated text in Calib.txt									
Element	Median_inmax_uA	Median_tau_s	Iso_fraction	f_ox	uC_exponential	uC_inmax	i_CP_adjust_uA	Test Region	Test Points/V
Fe	138243	79.63	400	0.99750	10981067	3018933	346	Top	22.2898182
Fe	138243	79.63	400	0.99750	10981067	3018933	346	Bottom	22.2898182
Cal & Ratio									
Duration/s	nth Test	Top	Bottom	mA	Test Points/V	Test Points/V	Composite Cal Factor/	Precision Ratio	Weight Factors
100	1	138.24	124.45	100.000	5.000	1.0000322	0.9864092	2.156%	22.4
100	2	138.24	124.45	100.000	5.000	1.0000271	0.9864103	2.156%	22.4
100	3	138.24	124.45	100.000	5.000	1.0000240	0.9864161	2.156%	22.4
100	4	138.24	124.45	100.000	5.000	1.0000234	0.9864089	2.156%	22.3
100	5	138.24	124.45	100.000	5.000	1.0000324	0.9864237	2.156%	22.3
100	6	138.24	124.45	100.000	5.000	1.0000408	0.9864123	2.156%	22.3
100	7	138.24	124.45	100.000	5.000	1.0000483	0.9864120	2.156%	22.3
100	8	138.24	124.45	100.000	5.000	1.0000527	0.9864104	2.156%	22.2
100	9	138.24	124.45	100.000	5.000	1.0000555	0.9864042	2.156%	22.2
100	10	138.24	124.45	100.000	5.000	1.0000517	0.9864085	2.156%	22.2
100	11	138.24	124.45	100.000	5.000	1.0000491	0.9864098	2.156%	22.2
Precision									
Ratio	Weight	Cal Temp	Average Cal Factor/	Cal Set, RSD%	Test Point Ave.	Ratio RSD, %	Ratio FYI		
1.0000387	0.00117%	0.9864096	0.00066%						

Figure 4-1. Example of output to Calibration text file

4.2 Iron Blank/Sample Measurements

Blank/Sample Measurements were performed using Sulfuric acid solutions and Iron at SRNL using the SRNL Cell. This test helped identified areas where the software was not functioning as expected. If an

unexpected outcome was obtained, the software engineer was notified. The software engineer performed changes to the program, as necessary, and testing was resumed. This process was repeated until the software produced expected outcomes. The testing progressed to testing with plutonium.

4.3 Plutonium Blank/Sample Measurements

Blank/Sample Measurement using Sulfuric acid solutions and plutonium using the SRNL Cell were performed. Plutonium standards were measured with HT Basic and LabVIEW software. The instrument and cell assembly used were the same. The only difference in the measurements is the software. All results with LabVIEW are within the laboratory limits.

Table 4-1 summarizes the data obtain. A t-test indicated the difference in the data is not statistically significant.

Table 4-1. Summary of plutonium measurements.

Software	Number of measurements	Average %RD (Pu corrected for Iron)	Std Deviation
HT Basic	10	-0.055	0.036
LabVIEW	10	0.050	0.067

4.4 Creation of files for data archiving

It was confirmed that the Data folder for the new software contained the required files. Table 4-2 summarizes the most important files. Figure 4-2 captures all the files generated by the software.

Table 4-2. Main files generated by software.

Filename	File content	Equivalent file in HT Basic
Blank.CXL	CXL file for the blank that contains the data collected during analysis	Blank.CXL
DateTime.CXL	CXL file for each ample analyzed with date and time stamp that contains the data collected during analysis	DateTime.CXL
Calib	Calibration txt file, results of continuous calibration	Calib
logbook	Text file that contains in chronological order blanks, samples, formal potentials	logbook
Sample	Text file that contains in chronological order blanks, samples, formal potentials with more detail information	samples

Name	Date modified	Type	Size
22May091103	5/9/2022 12:55 PM	CXL File	9 KB
22May091341	5/9/2022 2:15 PM	CXL File	10 KB
Blank	5/9/2022 2:36 PM	Microsoft Excel 97...	15 KB
Blankpreox	5/9/2022 10:14 AM	CXL File	12 KB
CALFACTR.CAL	5/9/2022 9:24 AM	CAL File	1 KB
Calib	5/9/2022 9:24 AM	Text Document	5,809 KB
charging	5/9/2022 2:34 PM	Text Document	14 KB
CompositeCals	5/9/2022 9:24 AM	Text Document	80 KB
Errors.lbk	5/9/2022 10:13 AM	LBK File	23 KB
Errors	5/9/2022 10:13 AM	Text Document	23 KB
LASTBLNK	5/9/2022 2:48 PM	Text Document	1 KB
logbook	5/10/2022 7:19 AM	Text Document	2 KB
postchg	5/9/2022 2:34 PM	Text Document	1 KB
Print1.bak	5/9/2022 1:17 PM	BAK File	5 KB
Print1	5/9/2022 2:38 PM	Text Document	5 KB
recdnum	5/9/2022 2:01 PM	Text Document	1 KB
Sample	5/9/2022 2:38 PM	Text Document	79 KB
Spl_i100	5/9/2022 1:54 PM	Text Document	15 KB
Spl_i100	5/9/2022 1:54 PM	Microsoft Excel 97...	12 KB
Spl_i101	5/9/2022 11:15 AM	Microsoft Excel 97...	12 KB

Figure 4-2. Files created by new LabVIEW program on Data folder.

4.5 Installation and operation at NML CPC

Before the installation of the new program at NML, the NML instrument was calibrated. See Appendix A for Calibration Report. The instrument was left overnight running continuous electrical calibrations. The next day a blank/sample determination was performed. Results were within the expected parameters. The formal potential was determined. All files for HT Basic were backed up.

A new computer was provided by IAEA personnel. The LabVIEW program was installed on IAEA NML computer. System parameters were adjusted for the instrument and cell configuration. The program was left running continuous electrical calibrations over lunch. Blank/Sample Measurement using Sulfuric acid solutions and plutonium using the IAEA NML cell were performed. Small adjustments to the program were performed. Final results displayed are total mg Pu. Results are within laboratory limit and agree with previous results. The software is functioning as expected and producing expected results, comparable to the HT Basic Software.

The IAEA NML personnel were trained on software use.

5.0 Conclusions

The IAEA NML CPC instrument was calibrated. The software was updated to LabVIEW. The software is functioning as expected and producing expected results, comparable to the HT Basic Software. Personnel were trained on the use of the new software.

6.0 Recommendations, Path Forward or Future Work

It is recommended that the personnel continue monitoring the performance of the new software and report any unexpected outcomes to SRNL personnel for evaluation.

Appendix A. Calibration Report

IAEA New Coulometer System, Calibrated on 2022-APR-25				
Lab Temperature 20.8 °C				
Aglient 34980A System Voltmeter Calibration				
Cross Check 34401A	Agilent 34401A / V	Fluke 87	Relative Diff, %	
IAEA Fluke Voltage Source				
0	0.0000700	0.0001000		
0.5	0.5000500	0.5000000	0.0100%	
1	1.0000300	1.0000000	0.0030%	
1.5	1.5000000	1.5000000	0.0000%	
2	2.0000000	2.0000000	0.0000%	
Cross Check 100 Ohm ASL Resistors at 22.6 Deg C (Use 23 Deg from Cal Report)				
34401A Four Wire Measurement	99.9980000	SRNL Cal Report	99.9982	0.00023
Condition of System prior to aligning the Digital Integrator				
K17 Relay Energized				
DAC 3001 = 2.5 V				
DAC 3002 = 1.0 V				
As Found Load Impedance and System Calibration Resistor Measurement				
R _{Std} resistor				
SRNL Calibration	Temp / °C	Resistance / Ω		
	20	99.998462		
	23	99.998225		
Temperature coefficient / Ω °C ⁻¹			-0.00008	
Actual Temperature / °C	21			
Temperature Δ from 20 °C	1.00000			
Change in Resistance / ΔΩ		-0.00008		
Corrected Resistance / Ω		99.99838		
R_{Load} Measurement				
V _{Std} / V	0.994520			
V _{Load} / V	0.199750			
I _{Load} / A	0.0099454		previous	% shift
R _{Load} / Ω	20.08474		20.08437	0.00185

Appendix A. Calibration Report, continuation

R_{Cal} Measurement				
V _{Cal} / V	0.99034			
V _{Load} / V	0.39787			
I _{Cal} / A	0.019810		previous	% shift
R _{Cal} / Ω	49.9930		49.996000	-0.00596
Entered new R _{Load} and R _{Cal} into CPCInstr.par file on:		49.9929	(Entered at IAEA Using Calculator)	
Conditions of System prior to aligning the Digital Integrator				
DAC 3001 = 2.5 V				
DAC 3002 = see chart				
K17 Relay Energized				
Voltage-to-Frequency Converter Alignment Check 09-14-2016, "As Found" same as "As left", i.e., no change				
DVM across load = 0.072e-3 V				
	As Found		As Left	
Readout	999.8	Hz	1000.0	Hz
Offset	1000.1	Hz	1000.1	Hz
Clock	10,000.11	Hz	10,000.11	Hz
34401A Load Voltage / V	HP53132A / Hz	With offset subtracted / Hz	Theoretical (without offset) / Hz	% Non Linearity
As Found				
2.0000	20996.7	19996.900	20000.000	-0.015
As Left				
2.00000	21000.3	20000.200	20000.000	0.001

Savannah River National Laboratory Measurement & Test Equipment			
1) ASL Standard Resistor Model RR100, SRNL M&TE# 54762, Expires 10/14/2022			
2) Agilent 34401A Digital Volt Meter, SRNL M&TE# 49750, Expires 2/3/23			
3) Agilent 53131A Universal Counter, SRNL M&TE# 53520, Expires 7/29/2022			
4) Omega HH41 Digital Thermometer, M&TE#49662, Exp 4/4/2023			
5) Fluke 87, M&TE #510137, Exp 3/10/2023			

Distribution:

Steven Balsley, Safeguards Analytical Services Director, Department of Safeguards, International Atomic Energy Agency

Mika Sumi, Laboratory Head, Nuclear Material Laboratory, Office of Safeguards Analytical Services| Department of Safeguards, International Atomic Energy Agency

Geoges Duhamel, Chemistry Team leader, Nuclear Material Laboratory, Department of Safeguards, International Atomic Energy Agency

Markus Kohl, Chemist, Nuclear Material Laboratory, Department of Safeguards, International Atomic Energy Agency