

**Contract No:**

This document was prepared in conjunction with work accomplished under Contract No. DE-AC09-08SR22470 with the U.S. Department of Energy (DOE) Office of Environmental Management (EM).

**Disclaimer:**

This work was prepared under an agreement with and funded by the U.S. Government. Neither the U. S. Government or its employees, nor any of its contractors, subcontractors or their employees, makes any express or implied:

- 1 ) warranty or assumes any legal liability for the accuracy, completeness, or for the use or results of such use of any information, product, or process disclosed; or
- 2 ) representation that such use or results of such use would not infringe privately owned rights; or
- 3) endorsement or recommendation of any specifically identified commercial product, process, or service.

Any views and opinions of authors expressed in this work do not necessarily state or reflect those of the United States Government, or its contractors, or subcontractors.

Subject: SRNL Coulometer for LANL - SRNL-Testing 2020

Authors: J. V. Cordaro, M. E. Morales-Arteaga

Scope:

SRNL fabricated two coulometers that have been delivered to LANL. All testing for electrical components is performed by R&D engineering and the results are attached to this report. Instrument testing for the first instrument was completed on August 2018 and installation at LANL was completed February 2019. This report addresses testing performed at FH Analytical Laboratory for the second instrument that arrived at LANL on June 5th, 2020. The testing is performed using Plutonium working standards. The instrument was kept in a radiologically clean area. The component calibration of the second LANL coulometer was completed in SRNL Building 723-A.

Summary:

The instrument was moved to FH analytical Laboratory and left for several days to perform electrical calibrations. The evaluation of electrical calibrations and other factors presented in more detail indicate that the performance of the instrument matches the typical performance for SRNL Coulometers.

Plutonium working standard measurements were also performed, a total of 10 measurements were performed using the SRNL cell setup and SRNL Coulometer for LANL. A total of 10 measurements were also performed with SRNL cell setup and SRNL Coulometer. A Grubs test was performed and indicated an outlier on each set of measurements. The outlier was not included in calculations. The sample aliquots each contained nominally 15 mg Pu and were measured using the SRNL method (ISO 12182, ASTM C1165).

Instrument	Average Recovery	RSD	%RD	Number of determinations
LANL-2	100.0057	0.0005	0.05%	10
SRNL	100.0962	0.0005	0.05%	10

Table 1. Summary of results for LANL instrument testing (does not include outliers)

Results:

Electrical calibration data generated at the FH Laboratory before the SRNL coulometer system was delivery to LANL were evaluated.

Precision estimates calculated from sets of routine electrical calibration results were a maximum of 0.0016% RSD, 1-sigma, when laboratory temperatures were stable. A relative standard deviation of 0.0040% RSD was found for a total of 650 values. This performance matches the typical performance for SRNL Coulometers.




Figure 1. Results of Calibration Factor over time.

The integrator response test data, the comparison of the 20-Ohm load resistor versus the 50-Ohm electrical calibration resistor data, and the integrator linearity test from 100 uA to 150 mA data, were all converted into graphs for ease of inspection. The graphs all have similar features. These tests were performed separately, but they are interrelated because the components being measured are interconnected in the potentiostat and integrator circuits. A typical set of graphs are shown below for linearity testing, integrator response testing, and load resistor testing:




Figure 2. Typical results from linearity testing




Figure 3. Typical results for Integrator Response testing.




Figure 4. Typical results for load resistor testing

The graph of the Load resistor vs. Electrical calibration resistor is expected to be a mirror image of the integrator response and electrical calibration graphs.

In all three tests, the constant current levels generated by the potentiostat circuit covered the range from 0.2 to 150 mA. In all test, this current flows through both the load and electrical calibration resistors. At higher currents these two resistors are heated more than at lower currents. The shape of the three curves at currents above 5 mA may be dictated more by the minor differences in the thermal properties of the 50-Ohm calibration resistance (two 100-Ohm resistors in parallel) versus the 20-Ohm load resistance (five 100-Ohm resistors in parallel) then by all other potential contributors combined. At the lower constant current levels, the shape of the curves from the three tests may be driven more by the systematic uncertainty in the measuring of the relatively small voltage drops across the load and calibration resistors at low currents than the actual non-linearity of the integration circuit.

A total of 10 Plutonium measurements were performed using the SRNL cell setup and SRNL Coulometer for LANL. Plutonium measurements were also performed with SRNL cell setup and SRNL Coulometer. A Grubbs test was performed and indicated an outlier on each set of measurements. The outlier was not included in calculations. An ANOVA single factor test was performed and indicated no statistical difference between groups. Results are presented in Table 2 and Table 3. The sample aliquots each contained nominally 15 mg Pu and were measured using the SRNL method (ISO 12183, ASTM C1165).

Analysis Date	% RD	Recovery, %
5/6/2020	0.0818	100.08
5/7/2020	0.0679	100.07
5/7/2020	0.0733	100.07
5/18/2020	-0.0128	99.99
5/18/2020	0.0044	100.00
5/19/2020	-0.0374	99.96
5/19/2020	-0.0419	99.96
5/19/2020	-0.0554	99.94
5/20/2020	-0.0026	100.00
5/20/2020	-0.0206	99.98

Average:	0.0057	100.0057
Std Deviation:	0.0005	0.0005

Table 2. Summary of results using LANL coulometer.

Based upon these test results the system was assessed by SRNL staff as having met the performance and design criteria used to determine proper functionality of the system.

Attached to this report are the following:

- Attachment 1. Component Calibration Results
- Attachment 2. Calib file documenting the calibration factor
- Attachment 3. Linear file documenting the linearity of the instrument
- Attachment 4. Instrument Par file that documents the value of the load impedance and calibration resistor.

## Attachment 1. Component Calibration Results

LANL3 New Coulometer System, Calibrated on 2020-04-30-2020				
Lab Temperature 21 °C				
Agilent 34980A System Voltmeter Calibration				
HP3458 Std	Mux 2001	Mux 2003	Slope	Diff Percent
	Cal Resistor Volts	Load Volts		
2.4805000	2.4804900		0.9992600	0.000
0.9969400		0.996950	0.9938600	-0.001
Condition of System prior to aligning the Digital Integrator				
K17 Relay Energized				
DAC 3001 = 2.5 V				
DAC 3002 = 2.5 V				
As Found Load Impedance and System Calibration Resistor Measurement				
R <sub>std</sub> resistor				
SRNL Calibration	Temp / °C	Resistance / Ω		
	20	99.998448		
	23	99.998211		
Temperature coefficient / Ω °C <sup>-1</sup>			-0.00008	
Actual Temperature / °C	21			
Temperature Δ from 20 °C	1.00000			
Change in Resistance / ΔΩ		-0.00008		
Corrected Resistance / Ω		99.99837		

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

<b>R<sub>Load</sub> Measurement</b>				
V <sub>Std</sub> / V	2.486240			
V <sub>Load</sub> / V	0.499640			
I <sub>Load</sub> / A	0.0248628			
R <sub>Load</sub> / Ω	<b>20.0959</b>			
<b>R<sub>Cal</sub> Measurement</b>				
V <sub>Cal</sub> / V	2.48062			
V <sub>Load</sub> / V	0.996980			
I <sub>Cal</sub> / A	0.0496112			
R <sub>Cal</sub> / Ω	<b>50.0013</b>			
Entered new R <sub>Load</sub> and R <sub>Cal</sub> into CPCInstr.par file on:	4/30/2020			

Voltage-to-Frequency Convertor Alignment Check 04-30-2020				
Readout	999.8	Hz	999.8	
Offset	1000.5	Hz	1000.5	
Clock	9,999.95	Hz		

3458 Load Voltage / V	HP53132A / Hz	With offset subtracted / Hz	Theoretical (without offset) / Hz	% Non Linearity
0.0000	999.8	0.000	0.000	
0.5	6000.2	5000.400	5000.000	0.008
1.00000	11000.6	10000.800	10000.000	0.008
2.00000	21000.5	20000.700	20000.000	0.004

Savannah River National Laboratory Measurement & Test Equipment

- 1) ASL Standard Resistor Model RR100, SRNL M&TE# 3-3110, Exp 06/26/2020  
At 20 °C,  $R_{Std} = 99.9984482 \Omega$ ; At 23 °C,  $R_{Std} = 99.998211 \Omega$  At 22Deg C 99.99829 Ω

2) Agilent 3458 Digital Volt Meter, SRNL M&TE# 3-4680, Exp 5/9/2020

3) Agilent 3458 Digital Volt Meter, SRNL M&TE# 3-6103, Exp 8/14/2020

4) Agilent 53132A Universal Counter, SRNL M&TE# 3-6106, Expires 4/29/2020  
Cross Checked performed to validate performance, calibration expired less than 24 hours before cal expired due to COVID-19

5) Temperature Sensor, M&TE# 3-4430, Exp 8/20/2020

Attachment 2. Calib file documenting the calibration factor

\*\*\*\*\*

\*\*\*\*\*

JVC COMPONENT CALIBRATION OF LANL3 INSTRUMEMNT 04-30-2020

\*\*\*\*\*

\*\*\*\*\*

T&E Option 3

Date for Cal: 30 Apr 2020

Time for Cal: 12:44:06 Duration= 50 seconds  
.99997 .99998 .99999 .99999 .99999

1.00000 .99999 1.00000 .99999 1.00000

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999990

RSD: 0.0008 % AT 13:03:21 ON 30 Apr 2020

.99999 .99999 .99999 .99999 1.00000

.99999 .99999 .99999 .99999 .99999

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999993

RSD: 0.0002 % AT 13:22:36 ON 30 Apr 2020

.99999 .99999 .99999 .99999 .99999

.99999 1.00000 .99999 .99999 .99999

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999992

RSD: 0.0003 % AT 13:41:51 ON 30 Apr 2020

.99999 .99999 .99999

.99987 .99987 .99988 .99988 .99989

.99989 .99989 .99989 .99989 .99990

LENGTH OF TIME FOR: CALIBRATION= 200.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999885

RSD: 0.0009 % AT 18:21:24 ON 30 Apr 2020

.99990 .99990 .99990 .99991 .99991

.99991 .99990 .99990 .99990 .99989

LENGTH OF TIME FOR: CALIBRATION= 225.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999900

RSD: 0.0005 % AT 19:39:53 ON 30 Apr 2020

.99989 .99989 .99988 .99988 .99988

.99988 .99988 .99988 .99988 .99988

LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999881

RSD: 0.0004 % AT 21:06:17 ON 30 Apr 2020

.99988 .99988 .99988 .99988 .99988

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99988        .99988        .99988        .99988        .99988  
LENGTH OF TIME FOR: CALIBRATION=                          275.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999880  
RSD: 0.0002 %        AT 22:40:34 ON 30 Apr 2020

.99988        .99988        .99989        .99988        .99989  
.99989        .99989        .99989        .99989        .99989  
LENGTH OF TIME FOR: CALIBRATION=                          300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999886  
RSD: 0.0002 %        AT 00:23:45 ON 1 May 2020

.99989        .99989        .99989        .99989        .99990  
.99990        .99990        .99990        .99990        .99990  
LENGTH OF TIME FOR: CALIBRATION=                          325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999896  
RSD: 0.0003 %        AT 02:15:49 ON 1 May 2020

.99990        .99990        .99990        .99990        .99990  
.99990        .99990        .99990        .99990        .99990  
LENGTH OF TIME FOR: CALIBRATION=                          350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999902  
RSD: 0.0001 %        AT 04:15:46 ON 1 May 2020

.99990        .99990        .99990        .99990        .99990  
.99990        .99991        .99991        .99991        .99991  
LENGTH OF TIME FOR: CALIBRATION=                          375.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999905  
RSD: 0.0004 %        AT 06:23:38 ON 1 May 2020

.99991        .99991        .99991        .99991        .99991  
.99991        .99991        .99991        .99991        .99991  
LENGTH OF TIME FOR: CALIBRATION=                          400.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999911  
RSD: 0.0002 %        AT 08:40:23 ON 1 May 2020

.99991        .99991        .99990        .99990        .99990  
.99990        .99990        .99990        .99990        .99990  
LENGTH OF TIME FOR: CALIBRATION=                          425.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999901  
RSD: 0.0003 %        AT 11:05:02 ON 1 May 2020

.99989        .99989        .99989        .99989        .99988  
.99988        .99988        .99987        .99986        .99986  
LENGTH OF TIME FOR: CALIBRATION=                          450.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:                          .999879  
RSD: 0.0013 %        AT 13:38:34 ON 1 May 2020

.99985        .99985        .99985        .99985        .99985  
.99985        .99985        .99984        .99984        .99984  
LENGTH OF TIME FOR: CALIBRATION=                          475.0 SEC.

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

723A AVERAGE CALIBRATION FACTOR: .999846  
RSD: 0.0004 % AT 16:19:59 ON 1 May 2020

.99984 .99983 .99983 .99983 .99983  
.99982 .99982 .99982 .99983 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999829  
RSD: 0.0006 % AT 19:09:19 ON 1 May 2020

.99985 .99986 .99986 .99987 .99988  
.99989 .99989 .99988 .99988 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999873  
RSD: 0.0013 % AT 21:58:39 ON 1 May 2020

.99987 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999861  
RSD: 0.0003 % AT 00:47:59 ON 2 May 2020

.99986 .99986 .99986 .99987 .99987  
.99987 .99987 .99987 .99987 .99988  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999868  
RSD: 0.0004 % AT 03:37:18 ON 2 May 2020

.99988 .99988 .99988 .99988 .99988  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999883  
RSD: 0.0005 % AT 06:26:38 ON 2 May 2020

.99989 .99989 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999890  
RSD: 0.0002 % AT 09:15:58 ON 2 May 2020

.99988 .99988 .99988 .99987 .99987  
.99987 .99986 .99986 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999868  
RSD: 0.0012 % AT 12:05:17 ON 2 May 2020

.99985 .99984 .99984 .99983 .99983  
.99982 .99982 .99982 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999827  
RSD: 0.0013 % AT 14:54:37 ON 2 May 2020

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99981 .99981 .99982 .99983 .99983  
.99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999831  
RSD: 0.0012 % AT 17:43:57 ON 2 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999846  
RSD: 0.0003 % AT 20:33:17 ON 2 May 2020

.99986 .99986 .99986 .99987 .99987  
.99988 .99988 .99989 .99989 .99990  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999876  
RSD: 0.0013 % AT 23:22:36 ON 2 May 2020

.99989 .99988 .99988 .99987 .99986  
.99986 .99985 .99985 .99985 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999863  
RSD: 0.0016 % AT 02:11:56 ON 3 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999839  
RSD: 0.0001 % AT 05:01:16 ON 3 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999843  
RSD: 0.0002 % AT 07:50:36 ON 3 May 2020

.99985 .99985 .99984 .99984 .99984  
.99984 .99984 .99984 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999840  
RSD: 0.0005 % AT 10:39:56 ON 3 May 2020

.99983 .99982 .99982 .99982 .99983  
.99984 .99984 .99984 .99984 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999834  
RSD: 0.0010 % AT 13:29:15 ON 3 May 2020

.99985 .99985 .99985 .99985 .99984

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99984	.99984	.99984	.99983	.99983
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999840	
RSD:	0.0005 %	AT 16:18:35 ON	3 May 2020	
.99983	.99983	.99983	.99982	.99982
.99982	.99982	.99982	.99982	.99982
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999825	
RSD:	0.0003 %	AT 19:07:55 ON	3 May 2020	
.99982	.99982	.99983	.99983	.99983
.99983	.99984	.99984	.99985	.99985
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999834	
RSD:	0.0010 %	AT 21:57:15 ON	3 May 2020	
.99985	.99985	.99986	.99986	.99986
.99986	.99987	.99987	.99987	.99987
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999863	
RSD:	0.0007 %	AT 00:46:35 ON	4 May 2020	
.99988	.99988	.99988	.99988	.99989
.99989	.99989	.99989	.99988	.99987
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999884	
RSD:	0.0006 %	AT 03:35:54 ON	4 May 2020	
.99986	.99986	.99985	.99985	.99984
.99984	.99984	.99983	.99983	.99983
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999844	
RSD:	0.0010 %	AT 06:25:14 ON	4 May 2020	
.99983	.99983	.99983	.99982	.99982
.99982	.99982	.99982	.99982	.99982
.99982	.99982			
LENGTH OF TIME FOR: CALIBRATION=			200.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999822	
RSD:	0.0002 %	AT 18:39:10 ON	4 May 2020	
.99983	.99983	.99982	.99983	.99983
.99983	.99982	.99983	.99983	.99983
LENGTH OF TIME FOR: CALIBRATION=			225.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999825	
RSD:	0.0001 %	AT 19:57:40 ON	4 May 2020	
.99983	.99983	.99983	.99983	.99983
.99983	.99983	.99983	.99983	.99983

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999828  
RSD: 0.0002 % AT 21:24:03 ON 4 May 2020

.99983 .99983 .99983 .99983 .99983  
.99983 .99983 .99983 .99983 .99984  
LENGTH OF TIME FOR: CALIBRATION= 275.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0002 % AT 22:58:21 ON 4 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999838  
RSD: 0.0001 % AT 00:41:32 ON 5 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999842  
RSD: 0.0001 % AT 02:33:35 ON 5 May 2020

.99984 .99985 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999846  
RSD: 0.0001 % AT 04:33:33 ON 5 May 2020

.99985 .99985 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 375.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852  
RSD: 0.0002 % AT 06:41:25 ON 5 May 2020

.99986 .99986 .99986 .99986 .99986  
.99979 .99979 .99979 .99979 .99979  
.99979 .99979 .99979 .99979 .99979  
LENGTH OF TIME FOR: CALIBRATION= 200.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999791  
RSD: 0.0002 % AT 20:28:37 ON 5 May 2020

.99979 .99980 .99980 .99980 .99980  
.99980 .99980 .99980 .99980 .99980  
LENGTH OF TIME FOR: CALIBRATION= 225.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999797  
RSD: 0.0001 % AT 21:47:07 ON 5 May 2020

.99980 .99980 .99980 .99980 .99980  
.99980 .99980 .99980 .99980 .99980  
LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999802  
RSD: 0.0002 % AT 23:13:30 ON 5 May 2020

.99980 .99980 .99981 .99980 .99981  
.99981 .99981 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 275.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999805  
RSD: 0.0001 % AT 00:47:47 ON 6 May 2020

.99981 .99981 .99981 .99981 .99981  
.99981 .99981 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999809  
RSD: 0.0001 % AT 02:30:58 ON 6 May 2020

.99981 .99981 .99981 .99981 .99981  
.99981 .99981 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999812  
RSD: 0.0001 % AT 04:23:02 ON 6 May 2020

.99981 .99981 .99981 .99982 .99982  
.99982 .99982 .99982 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999817  
RSD: 0.0003 % AT 06:23:00 ON 6 May 2020

.99982 .99982 .99983 .99983 .99983  
.99983 .99983 .99982 .99983 .99983  
.99983 .99983 .99983 .99984 .99984  
.99984 LENGTH OF TIME FOR: CALIBRATION= 200.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0005 % AT 20:49:06 ON 6 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 225.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999844  
RSD: 0.0003 % AT 22:07:35 ON 6 May 2020

.99985 .99985 .99985 .99985 .99985  
.99986 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999855  
RSD: 0.0005 % AT 23:33:59 ON 6 May 2020

.99986 .99986 .99986 .99987 .99987  
.99987 .99987 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 275.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999868  
RSD: 0.0003 % AT 01:08:16 ON 7 May 2020

.99987 .99987 .99988 .99988 .99988  
.99988 .99988 .99988 .99988 .99988  
LENGTH OF TIME FOR: CALIBRATION= 300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999876  
RSD: 0.0002 % AT 02:51:27 ON 7 May 2020

.99988 .99988 .99988 .99988 .99988  
.99988 .99988 .99988 .99988 .99988  
LENGTH OF TIME FOR: CALIBRATION= 325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999880  
RSD: 0.0001 % AT 04:43:31 ON 7 May 2020

.99988 .99988 .99988 .99988 .99988  
.99988 .99988 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999883  
RSD: 0.0002 % AT 06:43:28 ON 7 May 2020

.99989 .99989 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 375.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999890  
RSD: 0.0001 % AT 08:51:20 ON 7 May 2020

.99989 .99985 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 200.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852  
RSD: 0.0003 % AT 20:08:57 ON 7 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 225.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999857  
RSD: 0.0001 % AT 21:27:26 ON 7 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999862  
RSD: 0.0002 % AT 22:53:50 ON 7 May 2020

.99987 .99987 .99987 .99987 .99987  
.99987 .99987 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 275.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999868

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0002 % AT 00:28:07 ON 8 May 2020

.99987 .99987 .99987 .99987 .99988  
.99987 .99988 .99988 .99988 .99988

LENGTH OF TIME FOR: CALIBRATION= 300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999875

RSD: 0.0002 % AT 02:11:18 ON 8 May 2020

.99988 .99988 .99988 .99988 .99988  
.99988 .99988 .99988 .99988 .99988

LENGTH OF TIME FOR: CALIBRATION= 325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999881

RSD: 0.0002 % AT 04:03:16 ON 8 May 2020

.99989 .99988 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989

LENGTH OF TIME FOR: CALIBRATION= 350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999887

RSD: 0.0002 % AT 06:03:13 ON 8 May 2020

.99989 .99989 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989

LENGTH OF TIME FOR: CALIBRATION= 375.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999893

RSD: 0.0001 % AT 08:11:05 ON 8 May 2020

.99990 .99989 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989

LENGTH OF TIME FOR: CALIBRATION= 400.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999892

RSD: 0.0003 % AT 10:27:50 ON 8 May 2020

.99988 .99988 .99988 .99988 .99988  
.99987 .99987 .99987 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 425.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999874

RSD: 0.0006 % AT 12:52:28 ON 8 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986

LENGTH OF TIME FOR: CALIBRATION= 450.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999862

RSD: 0.0001 % AT 15:26:00 ON 8 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986

LENGTH OF TIME FOR: CALIBRATION= 475.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999859

RSD: 0.0001 % AT 18:07:25 ON 8 May 2020

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99986	.99986	.99986	.99985	.99985
.99985	.99985	.99985	.99985	.99985
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999854	
RSD:	0.0002 %	AT 20:56:45 ON	8 May 2020	
.99985	.99985	.99985	.99986	.99986
.99986	.99986	.99986	.99986	.99986
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999856	
RSD:	0.0002 %	AT 23:46:04 ON	8 May 2020	
.99986	.99986	.99986	.99986	.99986
.99987	.99987	.99987	.99987	.99987
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999866	
RSD:	0.0003 %	AT 02:35:24 ON	9 May 2020	
.99987	.99987	.99987	.99987	.99988
.99988	.99988	.99988	.99988	.99989
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999877	
RSD:	0.0006 %	AT 05:24:43 ON	9 May 2020	
.99989	.99989	.99989	.99989	.99990
.99990	.99990	.99990	.99990	.99990
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999896	
RSD:	0.0003 %	AT 08:14:03 ON	9 May 2020	
.99990	.99990	.99990	.99990	.99990
.99990	.99990	.99990	.99989	.99990
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999897	
RSD:	0.0002 %	AT 11:03:23 ON	9 May 2020	
.99989	.99989	.99989	.99989	.99989
.99988	.99988	.99988	.99988	.99988
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999887	
RSD:	0.0005 %	AT 13:52:42 ON	9 May 2020	
.99988	.99988	.99988	.99988	.99987
.99987	.99987	.99987	.99987	.99987
LENGTH OF TIME FOR: CALIBRATION=			500.0 SEC.	
723A AVERAGE CALIBRATION FACTOR:			.999873	
RSD:	0.0004 %	AT 16:42:02 ON	9 May 2020	
.99987	.99987	.99987	.99987	.99987
.99987	.99986	.99986	.99986	.99986

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999865  
RSD: 0.0002 % AT 19:31:21 ON 9 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999864  
RSD: 0.0002 % AT 22:20:41 ON 9 May 2020

.99987 .99987 .99987 .99988 .99987  
.99988 .99988 .99988 .99988 .99988  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999876  
RSD: 0.0004 % AT 01:10:01 ON 10 May 2020

.99988 .99988 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999889  
RSD: 0.0004 % AT 03:59:20 ON 10 May 2020

.99989 .99989 .99990 .99990 .99990  
.99990 .99990 .99990 .99990 .99990  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999898  
RSD: 0.0003 % AT 06:48:40 ON 10 May 2020

.99990 .99990 .99990 .99990 .99990  
.99990 .99990 .99990 .99990 .99990  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999902  
RSD: 0.0002 % AT 09:38:00 ON 10 May 2020

.99989 .99989 .99989 .99989 .99988  
.99988 .99988 .99988 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999883  
RSD: 0.0008 % AT 12:27:19 ON 10 May 2020

.99987 .99987 .99987 .99987 .99987  
.99987 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999866  
RSD: 0.0003 % AT 15:16:39 ON 10 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999857

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0003 % AT 18:05:58 ON 10 May 2020

.99985 .99985 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852

RSD: 0.0001 % AT 20:55:18 ON 10 May 2020

.99985 .99985 .99985 .99985 .99985  
.99985 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999853

RSD: 0.0003 % AT 23:44:37 ON 10 May 2020

.99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999863

RSD: 0.0003 % AT 02:33:57 ON 11 May 2020

.99987 .99987 .99987 .99987 .99987  
.99987 .99988 .99988 .99988 .99988  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999874

RSD: 0.0003 % AT 05:23:16 ON 11 May 2020

.99988 .99988 .99988 .99988 .99988  
.99988 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999884

RSD: 0.0002 % AT 08:12:36 ON 11 May 2020

.99989 .99989 .99988 .99988 .99988  
.99988 .99987 .99987 .99987 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999877

RSD: 0.0009 % AT 11:01:56 ON 11 May 2020

.99986 .99986 .99985 .99985 .99985  
.99985 .99985 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999849

RSD: 0.0005 % AT 13:51:15 ON 11 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999839

RSD: 0.0003 % AT 16:40:35 ON 11 May 2020

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99984 .99983 .99983 .99983 .99983 .99983  
.99983 .99983 .99983 .99983 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0001 % AT 19:29:54 ON 11 May 2020

.99983 .99983 .99983 .99983 .99983 .99983  
.99983 .99983 .99983 .99983 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999831  
RSD: 0.0001 % AT 22:19:14 ON 11 May 2020

.99983 .99983 .99984 .99984 .99984 .99984  
.99984 .99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999838  
RSD: 0.0003 % AT 01:08:33 ON 12 May 2020

.99984 .99984 .99984 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999848  
RSD: 0.0003 % AT 03:57:53 ON 12 May 2020

.99985 .99985 .99985 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999857  
RSD: 0.0003 % AT 06:47:12 ON 12 May 2020

.99986 .99986 .99986 .99986 .99986 .99987  
.99987 .99986 .99986 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999864  
RSD: 0.0001 % AT 09:36:32 ON 12 May 2020

.99986 .99986 .99986 .99986 .99986 .99986  
.99986 .99986 .99986 .99986 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999862  
RSD: 0.0002 % AT 12:25:51 ON 12 May 2020

.99986 .99985 .99985 .99985 .99985 .99985  
.99985 .99984 .99984 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999848  
RSD: 0.0006 % AT 15:15:11 ON 12 May 2020

.99984 .99984 .99983 .99983 .99983 .99983  
.99983 .99983 .99983 .99983 .99983 .99983

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999831  
RSD: 0.0004 % AT 18:04:30 ON 12 May 2020

.99982 .99982 .99982 .99982 .99982  
.99982 .99982 .99982 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999822  
RSD: 0.0001 % AT 20:53:50 ON 12 May 2020

.99982 .99982 .99982 .99982 .99982  
.99983 .99982 .99982 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999824  
RSD: 0.0002 % AT 23:43:09 ON 12 May 2020

.99983 .99983 .99983 .99983 .99983  
.99983 .99983 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0003 % AT 02:32:29 ON 13 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999843  
RSD: 0.0004 % AT 05:21:49 ON 13 May 2020

.99985 .99985 .99985 .99985 .99985  
.99985 .99985 .99985 .99986 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852  
RSD: 0.0003 % AT 08:11:08 ON 13 May 2020

.99985 .99986 .99985 .99985 .99985  
.99985 .99985 .99985 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999850  
RSD: 0.0004 % AT 11:00:28 ON 13 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99983 .99983 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999837  
RSD: 0.0005 % AT 13:49:47 ON 13 May 2020

.99983 .99983 .99982 .99982 .99982  
.99982 .99982 .99982 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999820

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0004 % AT 16:39:07 ON 13 May 2020

.99981 .99981 .99981 .99981 .99981  
.99981 .99981 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999811

RSD: 0.0002 % AT 19:28:26 ON 13 May 2020

.99981 .99981 .99981 .99981 .99981  
.99981 .99981 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999807

RSD: 0.0001 % AT 22:17:46 ON 13 May 2020

.99981 .99981 .99981 .99981 .99981  
.99981 .99981 .99982 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999813

RSD: 0.0002 % AT 01:07:06 ON 14 May 2020

.99982 .99982 .99982 .99982 .99982  
.99982 .99982 .99982 .99982 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999822

RSD: 0.0003 % AT 03:56:25 ON 14 May 2020

.99983 .99983 .99983 .99983 .99983  
.99983 .99983 .99983 .99983 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999831

RSD: 0.0003 % AT 06:45:45 ON 14 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99984 .99987 .99987 .99987  
.99988 .99988 .99988 .99988 .99988  
.99988  
LENGTH OF TIME FOR: CALIBRATION= 200.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999876

RSD: 0.0005 % AT 21:44:53 ON 26 May 2020

.99988 .99988 .99988 .99988 .99989  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 225.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999886

RSD: 0.0002 % AT 23:03:38 ON 26 May 2020

.99989 .99989 .99989 .99989 .99989  
.99989 .99989 .99989 .99989 .99989  
LENGTH OF TIME FOR: CALIBRATION= 250.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999888

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0001 % AT 00:30:18 ON 27 May 2020

.99989 .99989 .99989 .99989 .99989  
.99989 .99990 .99990 .99989 .99990

LENGTH OF TIME FOR: CALIBRATION= 275.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999894

RSD: 0.0002 % AT 02:04:53 ON 27 May 2020

.99990 .99990 .99990 .99990 .99990  
.99990 .99990 .99990 .99990 .99990

LENGTH OF TIME FOR: CALIBRATION= 300.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999899

RSD: 0.0002 % AT 03:47:24 ON 27 May 2020

.99990 .99990 .99990 .99990 .99990  
.99990 .99990 .99990 .99990 .99990

LENGTH OF TIME FOR: CALIBRATION= 325.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999902

RSD: 0.0001 % AT 05:38:49 ON 27 May 2020

.99990 .99990 .99990 .99990 .99990  
.99990 .99990 .99990 .99990 .99991

LENGTH OF TIME FOR: CALIBRATION= 350.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999904

RSD: 0.0001 % AT 07:39:10 ON 27 May 2020

.99990 .99989 .99988 .99987 .99986  
.99986 .99985 .99985 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 375.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999867

RSD: 0.0018 % AT 09:47:26 ON 27 May 2020

.99985 .99986 .99986

T&E Option 3

Date for Cal: 29 May 2020

Time for Cal: 15:43:23 Duration= 50 seconds

.99992 .99990 .99989 .99988 .99987

.99986 .99986 .99985 .99985 .99984

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

Cal Temperatures:

30.6 30.6 30.6 30.6 30.6

30.6 30.6 30.6 30.6 30.6

723A AVERAGE CALIBRATION FACTOR: .999873

Temperature= 30.6

RSD: 0.0025 % AT 16:02:45 ON 29 May 2020

T&E Option 1  
Date for Cal: 29 May 2020  
Time for Cal: 16:03:12 Duration= 500 seconds  
.99981 .99979 .99978 .99977 .99977  
.99977 .99977 .99977 .99977 .99977  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:  
30.8 31.1 31.3 31.4 31.5  
31.5 31.6 31.5 31.5 31.5  
723A AVERAGE CALIBRATION FACTOR: .999779  
Temperature= 31.4  
RSD: 0.0012 % AT 18:53:09 ON 29 May 2020

.99978 .99978 .99978 .99978 .99979  
.99979 .99979 .99979 .99980 .99980  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:  
31.5 31.5 31.4 31.4 31.3  
31.2 31.2 31.1 31.0 30.9  
723A AVERAGE CALIBRATION FACTOR: .999787  
Temperature= 31.2  
RSD: 0.0009 % AT 21:43:07 ON 29 May 2020

.99980 .99980 .99981 .99981 .99981  
.99982 .99981 .99982 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:  
30.9 30.9 30.8 30.7 30.7  
30.6 30.7 30.7 30.6 30.6  
723A AVERAGE CALIBRATION FACTOR: .999812  
Temperature= 30.7  
RSD: 0.0006 % AT 00:33:05 ON 30 May 2020

.99982 .99982 .99983 .99983 .99983  
.99983 .99983 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:  
30.5 30.4 30.4 30.4 30.3  
30.3 30.2 30.2 30.2 30.2  
723A AVERAGE CALIBRATION FACTOR: .999831  
Temperature= 30.3  
RSD: 0.0006 % AT 03:23:03 ON 30 May 2020

.99984 .99984 .99985 .99985 .99985  
.99985 .99985 .99985 .99985 .99986  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.1    30.1    30.1    30.0    30.0  
30.0    29.9    29.9    29.8    29.8

723A AVERAGE CALIBRATION FACTOR:       .999849

Temperature= 30.0

RSD: 0.0005 %           AT 06:13:01 ON 30 May 2020

.99986               .99986               .99986               .99986               .99986  
.99986    .99986               .99986               .99986    .99986  
LENGTH OF TIME FOR: CALIBRATION=        500.0 SEC.

Cal Temperatures:

29.8    29.8    29.7    29.7    29.7  
29.6    29.6    29.6    29.6    29.6

723A AVERAGE CALIBRATION FACTOR:       .999861

Temperature= 29.7

RSD: 0.0002 %           AT 09:02:59 ON 30 May 2020

.99986               .99986               .99986               .99986               .99986  
.99986    .99985               .99985               .99985    .99985  
LENGTH OF TIME FOR: CALIBRATION=        500.0 SEC.

Cal Temperatures:

29.6    29.6    29.7    29.7    29.7  
29.7    29.8    29.8    29.9    29.9

723A AVERAGE CALIBRATION FACTOR:       .999855

Temperature= 29.7

RSD: 0.0005 %           AT 11:52:57 ON 30 May 2020

.99984               .99984               .99984               .99983               .99983  
.99983    .99982               .99982               .99982    .99982  
LENGTH OF TIME FOR: CALIBRATION=        500.0 SEC.

Cal Temperatures:

30.0    30.1    30.2    30.2    30.3  
30.3    30.4    30.5    30.5    30.6

723A AVERAGE CALIBRATION FACTOR:       .999829

Temperature= 30.3

RSD: 0.0008 %           AT 14:42:55 ON 30 May 2020

.99981               .99981               .99981               .99981               .99981  
.99980    .99980               .99980               .99980    .99980  
LENGTH OF TIME FOR: CALIBRATION=        500.0 SEC.

Cal Temperatures:

30.7    30.7    30.8    30.8    30.8  
30.8    30.8    30.9    31.0    31.0

723A AVERAGE CALIBRATION FACTOR:       .999806

Temperature= 30.8

RSD: 0.0005 % AT 17:32:52 ON 30 May 2020

.99979 .99979 .99979 .99979 .99979  
.99978 .99978 .99978 .99978 .99978  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

31.1 31.1 31.2 31.2 31.3  
31.2 31.3 31.3 31.3 31.3

723A AVERAGE CALIBRATION FACTOR: .999787

Temperature= 31.2

RSD: 0.0004 % AT 20:22:50 ON 30 May 2020

.99979 .99979 .99979 .99979 .99980  
.99980 .99980 .99981 .99981 .99981  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

31.3 31.2 31.2 31.1 31.0  
31.0 30.8 30.8 30.7 30.6

723A AVERAGE CALIBRATION FACTOR: .999798

Temperature= 31.0

RSD: 0.0010 % AT 23:12:48 ON 30 May 2020

.99982 .99982 .99982 .99982 .99982  
.99983 .99983 .99983 .99983 .99983  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.5 30.5 30.5 30.4 30.4  
30.3 30.3 30.2 30.2 30.2

723A AVERAGE CALIBRATION FACTOR: .999825

Temperature= 30.4

RSD: 0.0006 % AT 02:02:46 ON 31 May 2020

.99984 .99984 .99984 .99984 .99984  
.99984 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.1 30.1 30.0 30.0 30.0  
29.9 29.9 29.8 29.8 29.8

723A AVERAGE CALIBRATION FACTOR: .999845

Temperature= 29.9

RSD: 0.0006 % AT 04:52:43 ON 31 May 2020

.99986 .99986 .99986 .99986 .99986  
.99987 .99987 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

29.7 29.7 29.7 29.6 29.6  
29.6 29.5 29.4 29.4 29.4

723A AVERAGE CALIBRATION FACTOR: .999865

Temperature= 29.6

RSD: 0.0006 % AT 07:42:41 ON 31 May 2020

.99987 .99987 .99987 .99987 .99987  
.99987 .99987 .99987 .99987 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

29.4 29.4 29.4 29.4 29.4  
29.3 29.4 29.4 29.4 29.4

723A AVERAGE CALIBRATION FACTOR: .999871

Temperature= 29.4

RSD: 0.0001 % AT 10:32:40 ON 31 May 2020

.99987 .99986 .99986 .99986 .99986  
.99986 .99985 .99985 .99985 .99985  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

29.5 29.5 29.5 29.6 29.6  
29.7 29.7 29.8 29.8 29.9

723A AVERAGE CALIBRATION FACTOR: .999857

Temperature= 29.6

RSD: 0.0007 % AT 13:22:38 ON 31 May 2020

.99984 .99984 .99984 .99984 .99983  
.99983 .99983 .99983 .99982 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.0 30.0 30.1 30.1 30.2  
30.3 30.3 30.3 30.4 30.4

723A AVERAGE CALIBRATION FACTOR: .999832

Temperature= 30.2

RSD: 0.0007 % AT 16:12:35 ON 31 May 2020

.99982 .99982 .99982 .99981 .99981  
.99981 .99981 .99980 .99980 .99980  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.4 30.5 30.6 30.6 30.6  
30.7 30.8 30.8 30.8 30.8

723A AVERAGE CALIBRATION FACTOR: .999810

Temperature= 30.7

RSD: 0.0006 % AT 19:02:33 ON 31 May 2020

.99980 .99980 .99980 .99980 .99980  
.99981 .99981 .99981 .99981 .99982  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.8 30.8 30.8 30.8 30.8  
30.8 30.8 30.7 30.6 30.5

723A AVERAGE CALIBRATION FACTOR: .999806

Temperature= 30.7

RSD: 0.0005 % AT 21:52:31 ON 31 May 2020

.99982 .99982 .99983 .99983 .99983  
.99983 .99983 .99984 .99984 .99984  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

30.5 30.4 30.3 30.2 30.2  
30.1 30.2 30.1 30.1 30.0

723A AVERAGE CALIBRATION FACTOR: .999831

Temperature= 30.2

RSD: 0.0007 % AT 00:42:29 ON 1 Jun 2020

.99984 .99985 .99985 .99985 .99985  
.99986 .99986 .99986 .99986 .99987  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

29.9 29.9 29.8 29.8 29.7  
29.7 29.6 29.6 29.5 29.5

723A AVERAGE CALIBRATION FACTOR: .999856

Temperature= 29.7

RSD: 0.0007 % AT 03:32:27 ON 1 Jun 2020

.99987 .99987 .99987 .99987 .99988  
.99988 .99988 .99988 .99988 .99988  
LENGTH OF TIME FOR: CALIBRATION= 500.0 SEC.

Cal Temperatures:

29.5 29.4 29.4 29.4 29.3  
29.3 29.2 29.2 29.2 29.1

723A AVERAGE CALIBRATION FACTOR: .999877

Temperature= 29.3

RSD: 0.0006 % AT 06:22:24 ON 1 Jun 2020

.99989 .99989 .99989 .99990 .99990  
.99990 .99990 .99990 .99990 .99990

Attachment 3. Linear file documenting the linearity of the instrument

LANL3 DIGITAL INTEGRATOR RESPONSE TEST  
MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

			Calculated	Actual	Cal Factor
			Hz/V	Hz/V	uC/count
OFFSET RATE OF VFC1, Hz:	4972.0	999.2			
CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49797.0	10007.1	.99929
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49796.4	10007.0	.99930
CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49795.7	10006.9	.99931
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49783.3	10004.4	.99956
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49782.0	10004.1	.99959
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49775.0	10002.7	.99973
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49770.0	10001.7	.99983
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49768.7	10001.5	.99985
CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49766.0	10000.9	.99991
CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49765.7	10000.9	.99991
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49765.7	10000.9	.99991
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49765.4	10000.8	.99992
CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49766.2	10001.0	.99990
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990

CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990

Calibration Resistor, value in PAR, Ohms:	50.0012
Load Impedance, value in PAR, Ohms:	20.0959

#### 723A LOAD RESISTOR vs CALIBRATION RESISTOR

Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4426	ohms
1.7254 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2136	ohms
.5860 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1681	ohms
.3594 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1322	ohms
.1808 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1209	ohms
.1246 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1115	ohms
.0778 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1034	ohms
.0373 %				
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1009	ohms
.0249 %				
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.0997	ohms
.0192 %				
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0172 %				
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0148 %				
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0131 %				
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0115 %				
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0102 %				
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0094 %				
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0096 %				
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0097 %				
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0095 %				
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0977	ohms

.0092 %					
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms	
.0096 %					
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0977	ohms	
.0093 %					
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0972	ohms	
.0067 %					
CURRENT, mA:	65.0 mA	LOAD RESISTOR, ohms:	20.0971	ohms	
.0063 %					
CURRENT, mA:	70.0 mA	LOAD RESISTOR, ohms:	20.0972	ohms	
.0063 %					
CURRENT, mA:	75.0 mA	LOAD RESISTOR, ohms:	20.0972	ohms	
.0064 %					
CURRENT, mA:	80.0 mA	LOAD RESISTOR, ohms:	20.0972	ohms	
.0065 %					
CURRENT, mA:	85.0 mA	LOAD RESISTOR, ohms:	20.0972	ohms	
.0064 %					
CURRENT, mA:	90.0 mA	LOAD RESISTOR, ohms:	20.0973	ohms	
.0070 %					
CURRENT, mA:	95.0 mA	LOAD RESISTOR, ohms:	20.0974	ohms	
.0075 %					
CURRENT, mA:	100.0 mA	LOAD RESISTOR, ohms:	20.0975	ohms	
.0079 %					
CURRENT, mA:	100.2 mA	LOAD RESISTOR, ohms:	20.0976	ohms	
.0084 %					
CURRENT, mA:	100.6 mA	LOAD RESISTOR, ohms:	20.0977	ohms	
.0091 %					
CURRENT, mA:	101.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms	
.0094 %					
CURRENT, mA:	102.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms	
.0096 %					
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0979	ohms	
.0101 %					
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0980	ohms	
.0104 %					
CURRENT, mA:	110.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms	
.0110 %					
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0982	ohms	
.0116 %					
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0984	ohms	
.0125 %					
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms	
.0135 %					
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms	
.0146 %					
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms	
.0160 %					
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.0994	ohms	
.0174 %					
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.0997	ohms	

.0188 %  
CURRENT, mA: 150.0 mA LOAD RESISTOR, ohms: 20.1000 ohms  
.0205 %

723A MCP DIGITAL INTEGRATOR LINEARITY TEST  
LINEARITY TEST: 50mA

.99981 .99983 .99984 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999836

RSD: 0.0016 % AT 15:01:07 ON 30 Apr 2020

LINEARITY TEST: 99uA

.99958 .99618 .99760 .99685 .99767

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .997573

RSD: 0.1278 % AT 15:10:46 ON 30 Apr 2020

LINEARITY TEST: 298uA

.99866 .99901 .99898 .99857 .99879

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .998803

RSD: 0.0193 % AT 15:20:24 ON 30 Apr 2020

LINEARITY TEST: 496uA

.99904 .99922 .99918 .99914 .99943

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999200

RSD: 0.0143 % AT 15:30:03 ON 30 Apr 2020

LINEARITY TEST: 993uA

.99962 .99974 .99955 .99958 .99954

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.

723A AVERAGE CALIBRATION FACTOR: .999607

RSD: 0.0079 % AT 15:39:41 ON 30 Apr 2020

LINEARITY TEST: 3mA

.99977 .99977 .99980 .99981 .99980

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999793  
RSD: 0.0018 % AT 15:49:20 ON 30 Apr 2020

LINEARITY TEST: 5mA

.99984 .99987 .99984 .99987 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999853  
RSD: 0.0017 % AT 15:58:59 ON 30 Apr 2020

LINEARITY TEST: 10mA

.99987 .99987 .99987 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999871  
RSD: 0.0001 % AT 16:08:37 ON 30 Apr 2020

LINEARITY TEST: 30mA

.99988 .99988 .99988 .99988 .99987

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999878  
RSD: 0.0004 % AT 16:18:16 ON 30 Apr 2020

LINEARITY TEST: 50mA

.99988 .99988 .99987 .99988 .99988

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999879  
RSD: 0.0003 % AT 16:27:55 ON 30 Apr 2020

LINEARITY TEST: 50mA

.99989 .99989 .99989 .99988 .99989

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999889  
RSD: 0.0003 % AT 16:33:37 ON 30 Apr 2020

LINEARITY TEST: 69mA

.99989 .99989 .99989 .99988 .99989

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999886

RSD: 0.0003 % AT 16:39:18 ON 30 Apr 2020

LINEARITY TEST: 89mA

.99988 .99988 .99988 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999876  
RSD: 0.0004 % AT 16:45:00 ON 30 Apr 2020

LINEARITY TEST: 109mA

.99986 .99986 .99986 .99985 .99986

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999859  
RSD: 0.0004 % AT 16:50:42 ON 30 Apr 2020

LINEARITY TEST: 129mA

.99984 .99983 .99983 .99983 .99983

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0007 % AT 16:56:23 ON 30 Apr 2020

LINEARITY TEST: 149mA

.99981 .99980 .99979 .99979 .99979

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999797  
RSD: 0.0008 % AT 17:02:05 ON 30 Apr 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99984 .99985 .99986 .99986 .99987

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999858  
RSD: 0.0012 % AT 17:11:44 ON 30 Apr 2020

723A DIGITAL INTEGRATOR RESPONSE TEST

MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

Calculated Hz/V	Actual Hz/V	Cal Factor uC/count
--------------------	----------------	------------------------

OFFSET RATE OF VFC1, Hz: 4973.2 999.4

CURRENT, mA:	.2	INTEGRATOR	RESPONSE RATE:	49825.1	10012.8	.99872
CURRENT, mA:	.6	INTEGRATOR	RESPONSE RATE:	49797.6	10007.3	.99927
CURRENT, mA:	1.0	INTEGRATOR	RESPONSE RATE:	49797.9	10007.3	.99927
CURRENT, mA:	2.0	INTEGRATOR	RESPONSE RATE:	49784.6	10004.7	.99953
CURRENT, mA:	3.0	INTEGRATOR	RESPONSE RATE:	49777.6	10003.3	.99967
CURRENT, mA:	5.0	INTEGRATOR	RESPONSE RATE:	49773.8	10002.5	.99975
CURRENT, mA:	10.0	INTEGRATOR	RESPONSE RATE:	49771.3	10002.0	.99980
CURRENT, mA:	15.0	INTEGRATOR	RESPONSE RATE:	49769.0	10001.5	.99985
CURRENT, mA:	20.0	INTEGRATOR	RESPONSE RATE:	49768.0	10001.3	.99987
CURRENT, mA:	25.0	INTEGRATOR	RESPONSE RATE:	49767.9	10001.3	.99987
CURRENT, mA:	30.0	INTEGRATOR	RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	35.0	INTEGRATOR	RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	40.0	INTEGRATOR	RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	45.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	50.0	INTEGRATOR	RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	50.2	INTEGRATOR	RESPONSE RATE:	49766.3	10001.0	.99990
CURRENT, mA:	50.6	INTEGRATOR	RESPONSE RATE:	49766.0	10000.9	.99991
CURRENT, mA:	51.0	INTEGRATOR	RESPONSE RATE:	49766.3	10001.0	.99990
CURRENT, mA:	52.0	INTEGRATOR	RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	53.0	INTEGRATOR	RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	55.0	INTEGRATOR	RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	60.0	INTEGRATOR	RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	65.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	70.0	INTEGRATOR	RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	75.0	INTEGRATOR	RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	80.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	85.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	90.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	95.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	100.0	INTEGRATOR	RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	100.2	INTEGRATOR	RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	100.6	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	101.0	INTEGRATOR	RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	102.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	103.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	105.0	INTEGRATOR	RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	110.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	115.0	INTEGRATOR	RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	120.0	INTEGRATOR	RESPONSE RATE:	49767.2	10001.1	.99989
CURRENT, mA:	125.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	130.0	INTEGRATOR	RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	135.0	INTEGRATOR	RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	140.0	INTEGRATOR	RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	145.0	INTEGRATOR	RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	150.0	INTEGRATOR	RESPONSE RATE:	49767.1	10001.1	.99989

Calibration Resistor, value in PAR, Ohms: 50.0012  
Load Impedance, value in PAR, Ohms: 20.0959

723A LOAD RESISTOR vs CALIBRATION RESISTOR

Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4496	ohms
1.7603 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2203	ohms
.6192 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1696	ohms
.3670 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1328	ohms
.1836 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1220	ohms
.1300 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1120	ohms
.0801 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1045	ohms
.0431 %				
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1022	ohms
.0316 %				
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.1011	ohms
.0259 %				
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.1005	ohms
.0228 %				
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.0999	ohms
.0200 %				
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0184 %				
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0171 %				
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0157 %				
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0148 %				
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0152 %				
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0151 %				
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0153 %				
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0147 %				
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0149 %				
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0146 %				
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0983	ohms

.0122	%				
CURRENT, mA:	65.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0120	%				
CURRENT, mA:	70.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0118	%				
CURRENT, mA:	75.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0119	%				
CURRENT, mA:	80.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0119	%				
CURRENT, mA:	85.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0120	%				
CURRENT, mA:	90.0	mA	LOAD RESISTOR, ohms:	20.0984	ohms
.0124	%				
CURRENT, mA:	95.0	mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0131	%				
CURRENT, mA:	100.0	mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0133	%				
CURRENT, mA:	100.2	mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0137	%				
CURRENT, mA:	100.6	mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0144	%				
CURRENT, mA:	101.0	mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0147	%				
CURRENT, mA:	102.0	mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0150	%				
CURRENT, mA:	103.0	mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0151	%				
CURRENT, mA:	105.0	mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0155	%				
CURRENT, mA:	110.0	mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0160	%				
CURRENT, mA:	115.0	mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0168	%				
CURRENT, mA:	120.0	mA	LOAD RESISTOR, ohms:	20.0994	ohms
.0175	%				
CURRENT, mA:	125.0	mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0186	%				
CURRENT, mA:	130.0	mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0196	%				
CURRENT, mA:	135.0	mA	LOAD RESISTOR, ohms:	20.1001	ohms
.0210	%				
CURRENT, mA:	140.0	mA	LOAD RESISTOR, ohms:	20.1004	ohms
.0224	%				
CURRENT, mA:	145.0	mA	LOAD RESISTOR, ohms:	20.1007	ohms
.0238	%				
CURRENT, mA:	150.0	mA	LOAD RESISTOR, ohms:	20.1010	ohms
.0254	%				

LINEARITY TEST: 50mA

.99976        .99979        .99979        .99980        .99981

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999790  
RSD: 0.0017 %        AT 15:18:52 ON 4 May 2020

LINEARITY TEST: 97uA

.99741        .99684        .99729        .99730        .99489

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .996747  
RSD: 0.1064 %        AT 15:28:31 ON 4 May 2020

LINEARITY TEST: 297uA

.99919        .99853        .99839        .99873        .99892

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .998752  
RSD: 0.0317 %        AT 15:38:10 ON 4 May 2020

LINEARITY TEST: 495uA

.99942        .99891        .99924        .99908        .99942

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999213  
RSD: 0.0223 %        AT 15:47:48 ON 4 May 2020

LINEARITY TEST: 991uA

.99963        .99960        .99969        .99967        .99945

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999607  
RSD: 0.0096 %        AT 15:57:27 ON 4 May 2020

LINEARITY TEST: 3mA

.99975        .99979        .99978        .99974        .99978

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999768  
RSD: 0.0024 %        AT 16:07:06 ON 4 May 2020

LINEARITY TEST: 5mA

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99980            .99983            .99982            .99981            .99980

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999812  
RSD: 0.0012 %            AT 16:16:44 ON 4 May 2020

LINEARITY TEST: 10mA

.99984            .99984            .99983            .99983            .99984

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999838  
RSD: 0.0007 %            AT 16:26:23 ON 4 May 2020

LINEARITY TEST: 30mA

.99984            .99984            .99984            .99984            .99984

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999841  
RSD: 0.0002 %            AT 16:36:02 ON 4 May 2020

LINEARITY TEST: 50mA

.99985            .99984            .99984            .99984            .99984

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999843  
RSD: 0.0002 %            AT 16:45:40 ON 4 May 2020

LINEARITY TEST: 50mA

.99985            .99985            .99984            .99985            .99985

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999849  
RSD: 0.0004 %            AT 16:51:22 ON 4 May 2020

LINEARITY TEST: 69mA

.99985            .99985            .99984            .99984            .99985

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999846  
RSD: 0.0002 %            AT 16:57:04 ON 4 May 2020

LINEARITY TEST: 89mA

.99984            .99983            .99983            .99983            .99983

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999833  
RSD: 0.0005 % AT 17:02:46 ON 4 May 2020

LINEARITY TEST: 109mA

.99982 .99982 .99981 .99981 .99981

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999816  
RSD: 0.0007 % AT 17:08:28 ON 4 May 2020

LINEARITY TEST: 129mA

.99980 .99979 .99979 .99979 .99978

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999789  
RSD: 0.0006 % AT 17:14:09 ON 4 May 2020

LINEARITY TEST: 149mA

.99976 .99975 .99975 .99974 .99974

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999749  
RSD: 0.0008 % AT 17:19:51 ON 4 May 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99980 .99981 .99981 .99981 .99982

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999809  
RSD: 0.0009 % AT 17:29:30 ON 4 May 2020

723A DIGITAL INTEGRATOR RESPONSE TEST

MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

	Calculated Hz/V	Actual Hz/V	Cal Factor uC/count
--	--------------------	----------------	------------------------

OFFSET RATE OF VFC1, Hz: 4975.5 999.9

CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49759.7	9999.7	1.00003
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49782.6	10004.2	.99958

CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49768.2	10001.3	.99987
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49772.8	10002.3	.99977
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49767.9	10001.3	.99987
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49767.5	10001.2	.99988
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.2	.99988
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.1	.99989
CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.1	.99989
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989

Calibration Resistor, value in PAR, Ohms: 50.0012  
Load Impedance, value in PAR, Ohms: 20.0959

723A LOAD RESISTOR vs CALIBRATION RESISTOR  
Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4387	ohms
1.7058 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2097	ohms
.5664 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1653	ohms
.3456 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1335	ohms
.1874 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1203	ohms
.1215 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1107	ohms
.0738 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1044	ohms
.0425 %				
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1021	ohms
.0308 %				
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.1009	ohms
.0249 %				
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.1007	ohms
.0239 %				
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.1002	ohms
.0214 %				
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0194 %				
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0170 %				
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0165 %				
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0157 %				
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0159 %				
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0155 %				
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0150 %				
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0159 %				
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0148 %				
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0154 %				
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0987	ohms
.0141 %				
CURRENT, mA:	65.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0147 %				
CURRENT, mA:	70.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0147 %				
CURRENT, mA:	75.0 mA	LOAD RESISTOR, ohms:	20.0987	ohms

.0141 %				
CURRENT, mA:	80.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0145 %				
CURRENT, mA:	85.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0146 %				
CURRENT, mA:	90.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0148 %				
CURRENT, mA:	95.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0152 %				
CURRENT, mA:	100.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0157 %				
CURRENT, mA:	100.2 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0164 %				
CURRENT, mA:	100.6 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0165 %				
CURRENT, mA:	101.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0167 %				
CURRENT, mA:	102.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0171 %				
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0994	ohms
.0175 %				
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0995	ohms
.0178 %				
CURRENT, mA:	110.0 mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0184 %				
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0997	ohms
.0189 %				
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0999	ohms
.0198 %				
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.1001	ohms
.0208 %				
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.1003	ohms
.0218 %				
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.1005	ohms
.0231 %				
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.1008	ohms
.0244 %				
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.1011	ohms
.0260 %				
CURRENT, mA:	150.0 mA	LOAD RESISTOR, ohms:	20.1014	ohms
.0274 %				

723A MCP DIGITAL INTEGRATOR LINEARITY TEST  
LINEARITY TEST: 50mA

.99973            .99975            .99976            .99977            .99978

LENGTH OF TIME FOR: CALIBRATION=            50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:            .999758

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0017 % AT 17:08:20 ON 5 May 2020

LINEARITY TEST: 97uA

.99742 .99866 .99787 .99776 .99649

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .997640  
RSD: 0.0788 % AT 17:17:58 ON 5 May 2020

LINEARITY TEST: 296uA

.99855 .99881 .99904 .99850 .99865

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .998709  
RSD: 0.0219 % AT 17:27:37 ON 5 May 2020

LINEARITY TEST: 495uA

.99926 .99902 .99893 .99927 .99933

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999161  
RSD: 0.0176 % AT 17:37:16 ON 5 May 2020

LINEARITY TEST: 991uA

.99952 .99952 .99946 .99955 .99963

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999535  
RSD: 0.0064 % AT 17:46:54 ON 5 May 2020

LINEARITY TEST: 3mA

.99976 .99974 .99979 .99973 .99969

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999743  
RSD: 0.0037 % AT 17:56:33 ON 5 May 2020

LINEARITY TEST: 5mA

.99978 .99975 .99981 .99978 .99979

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999780  
RSD: 0.0021 % AT 18:06:12 ON 5 May 2020

LINEARITY TEST: 10mA

.99982        .99980        .99981        .99980        .99982

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999811  
RSD: 0.0009 %        AT 18:15:50 ON 5 May 2020

LINEARITY TEST: 30mA

.99981        .99980        .99981        .99981        .99981

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999808  
RSD: 0.0003 %        AT 18:25:29 ON 5 May 2020

LINEARITY TEST: 50mA

.99981        .99981        .99981        .99981        .99981

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999810  
RSD: 0.0002 %        AT 18:35:08 ON 5 May 2020

LINEARITY TEST: 50mA

.99982        .99982        .99981        .99981        .99982

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999816  
RSD: 0.0003 %        AT 18:40:49 ON 5 May 2020

LINEARITY TEST: 69mA

.99981        .99981        .99981        .99981        .99981

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999809  
RSD: 0.0003 %        AT 18:46:31 ON 5 May 2020

LINEARITY TEST: 89mA

.99980        .99980        .99980        .99980        .99980

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999801  
RSD: 0.0003 %        AT 18:52:13 ON 5 May 2020

LINEARITY TEST: 109mA

.99979            .99978            .99978            .99978            .99977

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999780  
RSD: 0.0004 %            AT 18:57:55 ON 5 May 2020

LINEARITY TEST: 129mA

.99976            .99976            .99975            .99975            .99975

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999755  
RSD: 0.0007 %            AT 19:03:36 ON 5 May 2020

LINEARITY TEST: 149mA

.99973            .99972            .99971            .99971            .99971

LENGTH OF TIME FOR: CALIBRATION=        25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999715  
RSD: 0.0010 %            AT 19:09:18 ON 5 May 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99976            .99977            .99978            .99978            .99979

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999776  
RSD: 0.0011 %            AT 19:18:57 ON 5 May 2020

723A DIGITAL INTEGRATOR RESPONSE TEST

MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

	Calculated Hz/V	Actual Hz/V	Cal Factor uC/count
--	--------------------	----------------	------------------------

OFFSET RATE OF VFC1, Hz:        4975.6        999.9

CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49750.1	9997.7	1.00023
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49735.0	9994.7	1.00053
CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49744.6	9996.6	1.00034
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49758.8	9999.5	1.00005
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49764.5	10000.6	.99994
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49768.5	10001.4	.99986
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49765.7	10000.9	.99991
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990

CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49766.1	10000.9	.99991
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.2	.99988
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989

Calibration Resistor, value in PAR, Ohms: 50.0012  
 Load Impedance, value in PAR, Ohms: 20.0959

723A LOAD RESISTOR vs CALIBRATION RESISTOR  
 Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR,ohms:	20.4181	ohms
1.6032 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR,ohms:	20.2139	ohms
.5875 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR,ohms:	20.1670	ohms

.3537 %					
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1326	ohms	
.1829 %					
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1198	ohms	
.1190 %					
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1118	ohms	
.0793 %					
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1043	ohms	
.0420 %					
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1021	ohms	
.0308 %					
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.1006	ohms	
.0235 %					
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.1004	ohms	
.0224 %					
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms	
.0194 %					
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0994	ohms	
.0174 %					
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms	
.0163 %					
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms	
.0150 %					
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms	
.0145 %					
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0988	ohms	
.0146 %					
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0987	ohms	
.0139 %					
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms	
.0144 %					
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms	
.0136 %					
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms	
.0137 %					
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms	
.0131 %					
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms	
.0134 %					
CURRENT, mA:	65.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms	
.0129 %					
CURRENT, mA:	70.0 mA	LOAD RESISTOR, ohms:	20.0984	ohms	
.0128 %					
CURRENT, mA:	75.0 mA	LOAD RESISTOR, ohms:	20.0984	ohms	
.0127 %					
CURRENT, mA:	80.0 mA	LOAD RESISTOR, ohms:	20.0984	ohms	
.0126 %					
CURRENT, mA:	85.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms	
.0128 %					
CURRENT, mA:	90.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms	

.0130 %				
CURRENT, mA:	95.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0133 %				
CURRENT, mA:	100.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0135 %				
CURRENT, mA:	100.2 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0143 %				
CURRENT, mA:	100.6 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0152 %				
CURRENT, mA:	101.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0149 %				
CURRENT, mA:	102.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0155 %				
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0155 %				
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0159 %				
CURRENT, mA:	110.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0164 %				
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0172 %				
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0995	ohms
.0179 %				
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0187 %				
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0196 %				
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.1002	ohms
.0213 %				
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.1004	ohms
.0225 %				
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.1007	ohms
.0239 %				
CURRENT, mA:	150.0 mA	LOAD RESISTOR, ohms:	20.1010	ohms
.0256 %				

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99976            .99978            .99979            .99979            .99980

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:        .999784

RSD: 0.0017 %            AT 17:28:48 ON 6 May 2020

LINEARITY TEST: 97uA

.99747            .99687            .99675            .99548            .99793

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .996901  
RSD: 0.0928 % AT 17:38:27 ON 6 May 2020

LINEARITY TEST: 296uA

.99805 .99894 .99965 .99945 .99916

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999050  
RSD: 0.0625 % AT 17:48:06 ON 6 May 2020

LINEARITY TEST: 494uA

.99921 .99955 .99953 .99931 .99917

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999351  
RSD: 0.0179 % AT 17:57:44 ON 6 May 2020

LINEARITY TEST: 991uA

.99966 .99963 .99955 .99940 .99964

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999574  
RSD: 0.0109 % AT 18:07:23 ON 6 May 2020

LINEARITY TEST: 3mA

.99974 .99977 .99976 .99975 .99976

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999755  
RSD: 0.0010 % AT 18:17:02 ON 6 May 2020

LINEARITY TEST: 5mA

.99982 .99981 .99981 .99981 .99980

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999808  
RSD: 0.0005 % AT 18:26:40 ON 6 May 2020

LINEARITY TEST: 10mA

.99985 .99984 .99985 .99983 .99984

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999841

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0009 % AT 18:36:19 ON 6 May 2020

LINEARITY TEST: 30mA

.99985 .99984 .99985 .99984 .99984

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999844  
RSD: 0.0003 % AT 18:45:58 ON 6 May 2020

LINEARITY TEST: 50mA

.99985 .99984 .99984 .99984 .99984

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999843  
RSD: 0.0001 % AT 18:55:36 ON 6 May 2020

LINEARITY TEST: 50mA

.99985 .99985 .99985 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852  
RSD: 0.0002 % AT 19:01:18 ON 6 May 2020

LINEARITY TEST: 69mA

.99985 .99984 .99984 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999847  
RSD: 0.0003 % AT 19:07:00 ON 6 May 2020

LINEARITY TEST: 89mA

.99984 .99984 .99983 .99983 .99984

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999837  
RSD: 0.0003 % AT 19:12:42 ON 6 May 2020

LINEARITY TEST: 109mA

.99982 .99982 .99982 .99982 .99982

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999820  
RSD: 0.0004 % AT 19:18:23 ON 6 May 2020

LINEARITY TEST: 129mA

.99980 .99979 .99979 .99979 .99979

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999791  
RSD: 0.0006 % AT 19:24:05 ON 6 May 2020

LINEARITY TEST: 149mA

.99976 .99975 .99975 .99975 .99974

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999752  
RSD: 0.0008 % AT 19:29:47 ON 6 May 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99980 .99981 .99982 .99982 .99983

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999816  
RSD: 0.0011 % AT 19:39:26 ON 6 May 2020

723A DIGITAL INTEGRATOR RESPONSE TEST

MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

	Calculated Hz/V	Actual Hz/V	Cal Factor uC/count
--	--------------------	----------------	------------------------

OFFSET RATE OF VFC1, Hz: 4975.0 999.8

CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49820.9	10011.9	.99881
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49783.1	10004.4	.99956
CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49764.6	10000.6	.99994
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49764.2	10000.6	.99994
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49769.8	10001.7	.99983
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49767.0	10001.1	.99989
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.0	.99990
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1	.99989
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989

CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49766.9	10001.1	.99989
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49766.8	10001.1	.99989
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49766.2	10001.0	.99990
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49766.2	10001.0	.99990
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49766.3	10001.0	.99990
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49766.2	10001.0	.99990
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0	.99990
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49766.1	10000.9	.99991
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49766.7	10001.1	.99989
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49766.3	10001.0	.99990
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0	.99990
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49766.6	10001.0	.99990
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49766.2	10001.0	.99990

Calibration Resistor, value in PAR, Ohms: 50.0012  
 Load Impedance, value in PAR, Ohms: 20.0959

#### 723A LOAD RESISTOR vs CALIBRATION RESISTOR Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4129	ohms
1.5777 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2160	ohms
.5975 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1686	ohms
.3616 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1315	ohms
.1773 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1193	ohms
.1164 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1105	ohms

.0729 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1037	ohms
.0389 %				
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1015	ohms
.0280 %				
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0195 %				
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0194 %				
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0159 %				
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0151 %				
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0984	ohms
.0127 %				
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0120 %				
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0118 %				
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0113 %				
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0122 %				
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0111 %				
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0111 %				
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0115 %				
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0980	ohms
.0106 %				
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0108 %				
CURRENT, mA:	65.0 mA	LOAD RESISTOR, ohms:	20.0980	ohms
.0106 %				
CURRENT, mA:	70.0 mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0101 %				
CURRENT, mA:	75.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0097 %				
CURRENT, mA:	80.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0096 %				
CURRENT, mA:	85.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0095 %				
CURRENT, mA:	90.0 mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0099 %				
CURRENT, mA:	95.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0108 %				
CURRENT, mA:	100.0 mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0114 %				
CURRENT, mA:	100.2 mA	LOAD RESISTOR, ohms:	20.0983	ohms

.0118 %				
CURRENT, mA:	100.6 mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0121 %				
CURRENT, mA:	101.0 mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0122 %				
CURRENT, mA:	102.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0128 %				
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0130 %				
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0139 %				
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0145 %				
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0153 %				
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0161 %				
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0171 %				
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0185 %				
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.0998	ohms
.0197 %				
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.1001	ohms
.0210 %				
CURRENT, mA:	150.0 mA	LOAD RESISTOR, ohms:	20.1004	ohms
.0226 %				

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99979            .99981            .99982            .99983            .99983

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:        .999817

RSD: 0.0015 %        AT 16:48:40 ON 7 May 2020

LINEARITY TEST: 96uA

.99838            .99692            .99630            .99724            .99678

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:        .997124

RSD: 0.0780 %        AT 16:58:18 ON 7 May 2020

LINEARITY TEST: 296uA

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99874            .99920            .99866            .99897            .99894

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .998902  
RSD: 0.0213 %            AT 17:07:57 ON 7 May 2020

LINEARITY TEST: 494uA

.99891            .99921            .99948            .99942            .99961

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999323  
RSD: 0.0275 %            AT 17:17:36 ON 7 May 2020

LINEARITY TEST: 991uA

.99962            .99951            .99947            .99969            .99949

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999557  
RSD: 0.0095 %            AT 17:27:14 ON 7 May 2020

LINEARITY TEST: 3mA

.99979            .99985            .99983            .99983            .99982

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999822  
RSD: 0.0022 %            AT 17:36:53 ON 7 May 2020

LINEARITY TEST: 5mA

.99984            .99984            .99986            .99985            .99985

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999848  
RSD: 0.0007 %            AT 17:46:32 ON 7 May 2020

LINEARITY TEST: 10mA

.99986            .99988            .99987            .99987            .99986

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:        .999871  
RSD: 0.0008 %            AT 17:56:10 ON 7 May 2020

LINEARITY TEST: 30mA

.99987            .99987            .99987            .99987            .99987

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999871  
RSD: 0.0003 % AT 18:05:49 ON 7 May 2020

LINEARITY TEST: 50mA

.99987 .99987 .99987 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999873  
RSD: 0.0001 % AT 18:15:28 ON 7 May 2020

LINEARITY TEST: 50mA

.99988 .99987 .99988 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999874  
RSD: 0.0002 % AT 18:21:09 ON 7 May 2020

LINEARITY TEST: 69mA

.99987 .99988 .99987 .99987 .99987

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999874  
RSD: 0.0002 % AT 18:26:51 ON 7 May 2020

LINEARITY TEST: 89mA

.99987 .99987 .99986 .99986 .99986

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999863  
RSD: 0.0005 % AT 18:32:33 ON 7 May 2020

LINEARITY TEST: 109mA

.99985 .99985 .99985 .99985 .99984

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999848  
RSD: 0.0006 % AT 18:38:15 ON 7 May 2020

LINEARITY TEST: 129mA

.99983 .99982 .99982 .99982 .99981

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999817

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0007 % AT 18:43:56 ON 7 May 2020

LINEARITY TEST: 149mA

.99979 .99978 .99978 .99978 .99977

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999780  
RSD: 0.0007 % AT 18:49:38 ON 7 May 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99982 .99983 .99984 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999838  
RSD: 0.0011 % AT 18:59:17 ON 7 May 2020

723A DIGITAL INTEGRATOR RESPONSE TEST

MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

		Calculated	Actual	Cal Factor
		Hz/V	Hz/V	uC/count
OFFSET RATE OF VFC1, Hz:	4974.5	999.7		
CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49717.8	9991.2 1.00088
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49747.1	9997.1 1.00029
CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49745.6	9996.8 1.00032
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49764.1	10000.5 .99995
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49769.2	10001.6 .99984
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49766.5	10001.0 .99990
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49766.3	10001.0 .99990
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49767.2	10001.2 .99988
CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3 .99987
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49767.1	10001.1 .99989
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.3 .99987
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3 .99987
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49767.9	10001.3 .99987
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49767.6	10001.2 .99988
CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3 .99987
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49767.9	10001.3 .99987
CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49767.7	10001.3 .99987
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2 .99988
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49767.5	10001.2 .99988
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3 .99987

CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49768.1	10001.3	.99987
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3	.99987
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.3	.99987
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49768.1	10001.3	.99987
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49767.6	10001.2	.99988
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49768.0	10001.3	.99987
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3	.99987
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3	.99987
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3	.99987
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49767.5	10001.2	.99988
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49767.5	10001.2	.99988
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49767.3	10001.2	.99988
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.3	.99987
CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49768.8	10001.5	.99985
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.3	.99987
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49767.7	10001.2	.99988
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49767.6	10001.2	.99988
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49767.4	10001.2	.99988

Calibration Resistor, value in PAR, Ohms:	50.0012
Load Impedance, value in PAR, Ohms:	20.0959

#### 723A LOAD RESISTOR vs CALIBRATION RESISTOR Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4532	ohms
1.7782 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2076	ohms
.5559 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1699	ohms
.3682 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1328	ohms
.1836 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1198	ohms
.1189 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1110	ohms
.0750 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1035	ohms
.0378 %				
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1014	ohms
.0275 %				
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.1002	ohms

.0215	%				
CURRENT, mA:	25.0	mA	LOAD RESISTOR, ohms:	20.0999	ohms
.0202	%				
CURRENT, mA:	30.0	mA	LOAD RESISTOR, ohms:	20.0991	ohms
.0163	%				
CURRENT, mA:	35.0	mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0148	%				
CURRENT, mA:	40.0	mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0132	%				
CURRENT, mA:	45.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0120	%				
CURRENT, mA:	50.0	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0117	%				
CURRENT, mA:	50.2	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0115	%				
CURRENT, mA:	50.6	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0114	%				
CURRENT, mA:	51.0	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0114	%				
CURRENT, mA:	52.0	mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0112	%				
CURRENT, mA:	53.0	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0115	%				
CURRENT, mA:	55.0	mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0111	%				
CURRENT, mA:	60.0	mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0103	%				
CURRENT, mA:	65.0	mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0098	%				
CURRENT, mA:	70.0	mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0101	%				
CURRENT, mA:	75.0	mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0095	%				
CURRENT, mA:	80.0	mA	LOAD RESISTOR, ohms:	20.0980	ohms
.0105	%				
CURRENT, mA:	85.0	mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0100	%				
CURRENT, mA:	90.0	mA	LOAD RESISTOR, ohms:	20.0980	ohms
.0106	%				
CURRENT, mA:	95.0	mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0110	%				
CURRENT, mA:	100.0	mA	LOAD RESISTOR, ohms:	20.0982	ohms
.0113	%				
CURRENT, mA:	100.2	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0119	%				
CURRENT, mA:	100.6	mA	LOAD RESISTOR, ohms:	20.0984	ohms
.0123	%				
CURRENT, mA:	101.0	mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0122	%				
CURRENT, mA:	102.0	mA	LOAD RESISTOR, ohms:	20.0985	ohms

.0131 %				
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms
.0130 %				
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0134 %				
CURRENT, mA:	110.0 mA	LOAD RESISTOR, ohms:	20.0987	ohms
.0139 %				
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0988	ohms
.0145 %				
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0990	ohms
.0156 %				
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0165 %				
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.0993	ohms
.0172 %				
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.0996	ohms
.0185 %				
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.0999	ohms
.0199 %				
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.1002	ohms
.0214 %				
CURRENT, mA:	150.0 mA	LOAD RESISTOR, ohms:	20.1005	ohms
.0232 %				

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99976                  .99978                  .99979                  .99979                  .99981

LENGTH OF TIME FOR: CALIBRATION=                  50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:                  .999788

RSD: 0.0018 %                  AT 12:44:54 ON 20 May 2020

LINEARITY TEST: 97uA

.99723                  .99746                  .99762                  .99683                  .99700

LENGTH OF TIME FOR: CALIBRATION=                  50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:                  .997228

RSD: 0.0325 %                  AT 12:54:35 ON 20 May 2020

LINEARITY TEST: 296uA

.99916                  .99885                  .99907                  .99860                  .99826

LENGTH OF TIME FOR: CALIBRATION=                  50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:                  .998786

RSD: 0.0366 %                  AT 13:04:15 ON 20 May 2020

LINEARITY TEST: 494uA

.99919 .99951 .99938 .99931 .99875

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999227  
RSD: 0.0291 % AT 13:13:56 ON 20 May 2020

LINEARITY TEST: 991uA

.99956 .99948 .99955 .99960 .99966

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999571  
RSD: 0.0069 % AT 13:23:37 ON 20 May 2020

LINEARITY TEST: 3mA

.99980 .99976 .99981 .99976 .99979

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999782  
RSD: 0.0022 % AT 13:33:17 ON 20 May 2020

LINEARITY TEST: 5mA

.99981 .99981 .99981 .99982 .99983

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999816  
RSD: 0.0011 % AT 13:42:58 ON 20 May 2020

LINEARITY TEST: 10mA

.99987 .99985 .99985 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999853  
RSD: 0.0011 % AT 13:52:38 ON 20 May 2020

LINEARITY TEST: 30mA

.99986 .99985 .99986 .99986 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999855  
RSD: 0.0003 % AT 14:02:19 ON 20 May 2020

LINEARITY TEST: 50mA

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

.99986            .99986            .99986            .99985            .99986

LENGTH OF TIME FOR: CALIBRATION=            50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:            .999857  
RSD: 0.0004 %            AT 14:12:00 ON 20 May 2020

LINEARITY TEST:        50mA

.99987            .99986            .99987            .99986            .99987

LENGTH OF TIME FOR: CALIBRATION=            25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:            .999866  
RSD: 0.0003 %            AT 14:17:42 ON 20 May 2020

LINEARITY TEST:        69mA

.99986            .99986            .99986            .99986            .99987

LENGTH OF TIME FOR: CALIBRATION=            25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR:            .999863  
RSD: 0.0003 %            AT 14:23:25 ON 20 May 2020

LINEARITY TEST:        89mA

723A DIGITAL INTEGRATOR RESPONSE TEST  
MEASURE Hz/VOLT AND THEORETICAL uCOULOMBS/COUNT AT SPECIFIED CURRENT LEVELS

		Calculated	Actual	Cal Factor
CURRENT, mA:		Hz/V	Hz/V	uC/count
OFFSET RATE OF VFC1, Hz:	4970.4	998.8		
CURRENT, mA:	.2	INTEGRATOR RESPONSE RATE:	49807.2	10009.2
CURRENT, mA:	.6	INTEGRATOR RESPONSE RATE:	49792.4	10006.2
CURRENT, mA:	1.0	INTEGRATOR RESPONSE RATE:	49789.5	10005.6
CURRENT, mA:	2.0	INTEGRATOR RESPONSE RATE:	49777.5	10003.2
CURRENT, mA:	3.0	INTEGRATOR RESPONSE RATE:	49775.0	10002.7
CURRENT, mA:	5.0	INTEGRATOR RESPONSE RATE:	49772.2	10002.2
CURRENT, mA:	10.0	INTEGRATOR RESPONSE RATE:	49767.8	10001.3
CURRENT, mA:	15.0	INTEGRATOR RESPONSE RATE:	49766.4	10001.0
CURRENT, mA:	20.0	INTEGRATOR RESPONSE RATE:	49766.1	10000.9
CURRENT, mA:	25.0	INTEGRATOR RESPONSE RATE:	49765.6	10000.8
CURRENT, mA:	30.0	INTEGRATOR RESPONSE RATE:	49765.0	10000.7
CURRENT, mA:	35.0	INTEGRATOR RESPONSE RATE:	49764.7	10000.6
CURRENT, mA:	40.0	INTEGRATOR RESPONSE RATE:	49764.8	10000.7
CURRENT, mA:	45.0	INTEGRATOR RESPONSE RATE:	49764.6	10000.6
CURRENT, mA:	50.0	INTEGRATOR RESPONSE RATE:	49764.7	10000.7
CURRENT, mA:	50.2	INTEGRATOR RESPONSE RATE:	49764.3	10000.6

CURRENT, mA:	50.6	INTEGRATOR RESPONSE RATE:	49764.3	10000.6	.99994
CURRENT, mA:	51.0	INTEGRATOR RESPONSE RATE:	49764.4	10000.6	.99994
CURRENT, mA:	52.0	INTEGRATOR RESPONSE RATE:	49764.4	10000.6	.99994
CURRENT, mA:	53.0	INTEGRATOR RESPONSE RATE:	49764.3	10000.6	.99994
CURRENT, mA:	55.0	INTEGRATOR RESPONSE RATE:	49764.4	10000.6	.99994
CURRENT, mA:	60.0	INTEGRATOR RESPONSE RATE:	49765.1	10000.7	.99993
CURRENT, mA:	65.0	INTEGRATOR RESPONSE RATE:	49765.3	10000.8	.99992
CURRENT, mA:	70.0	INTEGRATOR RESPONSE RATE:	49765.4	10000.8	.99992
CURRENT, mA:	75.0	INTEGRATOR RESPONSE RATE:	49765.3	10000.8	.99992
CURRENT, mA:	80.0	INTEGRATOR RESPONSE RATE:	49765.4	10000.8	.99992
CURRENT, mA:	85.0	INTEGRATOR RESPONSE RATE:	49765.5	10000.8	.99992
CURRENT, mA:	90.0	INTEGRATOR RESPONSE RATE:	49765.2	10000.7	.99993
CURRENT, mA:	95.0	INTEGRATOR RESPONSE RATE:	49765.3	10000.8	.99992
CURRENT, mA:	100.0	INTEGRATOR RESPONSE RATE:	49765.4	10000.8	.99992
CURRENT, mA:	100.2	INTEGRATOR RESPONSE RATE:	49765.4	10000.8	.99992
CURRENT, mA:	100.6	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	101.0	INTEGRATOR RESPONSE RATE:	49765.5	10000.8	.99992
CURRENT, mA:	102.0	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	103.0	INTEGRATOR RESPONSE RATE:	49765.7	10000.9	.99991
CURRENT, mA:	105.0	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	110.0	INTEGRATOR RESPONSE RATE:	49765.6	10000.8	.99992
CURRENT, mA:	115.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	120.0	INTEGRATOR RESPONSE RATE:	49765.5	10000.8	.99992
CURRENT, mA:	125.0	INTEGRATOR RESPONSE RATE:	49765.9	10000.9	.99991
CURRENT, mA:	130.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	135.0	INTEGRATOR RESPONSE RATE:	49765.9	10000.9	.99991
CURRENT, mA:	140.0	INTEGRATOR RESPONSE RATE:	49765.8	10000.9	.99991
CURRENT, mA:	145.0	INTEGRATOR RESPONSE RATE:	49765.7	10000.9	.99991
CURRENT, mA:	150.0	INTEGRATOR RESPONSE RATE:	49765.9	10000.9	.99991

Calibration Resistor, value in PAR, Ohms: 50.0012  
 Load Impedance, value in PAR, Ohms: 20.0959

723A LOAD RESISTOR vs CALIBRATION RESISTOR  
 Load Vs Cal measured with K17 Energized

CURRENT, mA:	.2 mA	LOAD RESISTOR, ohms:	20.4453	ohms
1.7386 %				
CURRENT, mA:	.6 mA	LOAD RESISTOR, ohms:	20.2159	ohms
.5971 %				
CURRENT, mA:	1.0 mA	LOAD RESISTOR, ohms:	20.1677	ohms
.3575 %				
CURRENT, mA:	2.0 mA	LOAD RESISTOR, ohms:	20.1312	ohms
.1757 %				
CURRENT, mA:	3.0 mA	LOAD RESISTOR, ohms:	20.1207	ohms
.1233 %				
CURRENT, mA:	5.0 mA	LOAD RESISTOR, ohms:	20.1101	ohms
.0708 %				
CURRENT, mA:	10.0 mA	LOAD RESISTOR, ohms:	20.1024	ohms

.0325 %					
CURRENT, mA:	15.0 mA	LOAD RESISTOR, ohms:	20.1003	ohms	
.0220 %					
CURRENT, mA:	20.0 mA	LOAD RESISTOR, ohms:	20.0991	ohms	
.0160 %					
CURRENT, mA:	25.0 mA	LOAD RESISTOR, ohms:	20.0985	ohms	
.0129 %					
CURRENT, mA:	30.0 mA	LOAD RESISTOR, ohms:	20.0980	ohms	
.0107 %					
CURRENT, mA:	35.0 mA	LOAD RESISTOR, ohms:	20.0976	ohms	
.0087 %					
CURRENT, mA:	40.0 mA	LOAD RESISTOR, ohms:	20.0973	ohms	
.0072 %					
CURRENT, mA:	45.0 mA	LOAD RESISTOR, ohms:	20.0971	ohms	
.0063 %					
CURRENT, mA:	50.0 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0055 %					
CURRENT, mA:	50.2 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0056 %					
CURRENT, mA:	50.6 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0057 %					
CURRENT, mA:	51.0 mA	LOAD RESISTOR, ohms:	20.0971	ohms	
.0058 %					
CURRENT, mA:	52.0 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0058 %					
CURRENT, mA:	53.0 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0058 %					
CURRENT, mA:	55.0 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0055 %					
CURRENT, mA:	60.0 mA	LOAD RESISTOR, ohms:	20.0966	ohms	
.0036 %					
CURRENT, mA:	65.0 mA	LOAD RESISTOR, ohms:	20.0965	ohms	
.0032 %					
CURRENT, mA:	70.0 mA	LOAD RESISTOR, ohms:	20.0965	ohms	
.0031 %					
CURRENT, mA:	75.0 mA	LOAD RESISTOR, ohms:	20.0965	ohms	
.0033 %					
CURRENT, mA:	80.0 mA	LOAD RESISTOR, ohms:	20.0966	ohms	
.0035 %					
CURRENT, mA:	85.0 mA	LOAD RESISTOR, ohms:	20.0967	ohms	
.0039 %					
CURRENT, mA:	90.0 mA	LOAD RESISTOR, ohms:	20.0967	ohms	
.0041 %					
CURRENT, mA:	95.0 mA	LOAD RESISTOR, ohms:	20.0968	ohms	
.0045 %					
CURRENT, mA:	100.0 mA	LOAD RESISTOR, ohms:	20.0970	ohms	
.0054 %					
CURRENT, mA:	100.2 mA	LOAD RESISTOR, ohms:	20.0971	ohms	
.0059 %					
CURRENT, mA:	100.6 mA	LOAD RESISTOR, ohms:	20.0971	ohms	

.0062 %				
CURRENT, mA:	101.0 mA	LOAD RESISTOR, ohms:	20.0973	ohms
.0069 %				
CURRENT, mA:	102.0 mA	LOAD RESISTOR, ohms:	20.0973	ohms
.0071 %				
CURRENT, mA:	103.0 mA	LOAD RESISTOR, ohms:	20.0974	ohms
.0075 %				
CURRENT, mA:	105.0 mA	LOAD RESISTOR, ohms:	20.0974	ohms
.0078 %				
CURRENT, mA:	110.0 mA	LOAD RESISTOR, ohms:	20.0976	ohms
.0084 %				
CURRENT, mA:	115.0 mA	LOAD RESISTOR, ohms:	20.0978	ohms
.0093 %				
CURRENT, mA:	120.0 mA	LOAD RESISTOR, ohms:	20.0979	ohms
.0101 %				
CURRENT, mA:	125.0 mA	LOAD RESISTOR, ohms:	20.0981	ohms
.0111 %				
CURRENT, mA:	130.0 mA	LOAD RESISTOR, ohms:	20.0983	ohms
.0121 %				
CURRENT, mA:	135.0 mA	LOAD RESISTOR, ohms:	20.0986	ohms
.0136 %				
CURRENT, mA:	140.0 mA	LOAD RESISTOR, ohms:	20.0989	ohms
.0150 %				
CURRENT, mA:	145.0 mA	LOAD RESISTOR, ohms:	20.0992	ohms
.0166 %				
CURRENT, mA:	150.0 mA	LOAD RESISTOR, ohms:	20.0995	ohms
.0181 %				

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99984            .99985            .99987            .99987            .99987

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:        .999860

RSD: 0.0014 %        AT 18:23:56 ON 26 May 2020

LINEARITY TEST: 98uA

.99664            .99822            .99607            .99557            .99643

LENGTH OF TIME FOR: CALIBRATION=        50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:        .996587

RSD: 0.1000 %        AT 18:33:37 ON 26 May 2020

LINEARITY TEST: 298uA

.99895            .99941            .99902            .99876            .99875

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .998980  
RSD: 0.0270 % AT 18:43:17 ON 26 May 2020

LINEARITY TEST: 496uA

.99925 .99942 .99921 .99932 .99934

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999309  
RSD: 0.0084 % AT 18:52:58 ON 26 May 2020

LINEARITY TEST: 992uA

.99955 .99964 .99975 .99966 .99963

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999645  
RSD: 0.0070 % AT 19:02:39 ON 26 May 2020

LINEARITY TEST: 3mA

.99979 .99977 .99983 .99974 .99981

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999788  
RSD: 0.0034 % AT 19:12:19 ON 26 May 2020

LINEARITY TEST: 5mA

.99986 .99986 .99984 .99985 .99985

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999852  
RSD: 0.0007 % AT 19:22:00 ON 26 May 2020

LINEARITY TEST: 10mA

.99989 .99989 .99988 .99989 .99990

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999888  
RSD: 0.0007 % AT 19:31:40 ON 26 May 2020

LINEARITY TEST: 30mA

.99989 .99989 .99989 .99989 .99989

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999889

SRNL Coulometer for LANL - SRNL-Testing 2020  
SRNL-L4610-2020-00009

RSD: 0.0003 % AT 19:41:21 ON 26 May 2020

LINEARITY TEST: 50mA

.99989 .99989 .99989 .99988 .99989

LENGTH OF TIME FOR: CALIBRATION= 50.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999888  
RSD: 0.0003 % AT 19:51:02 ON 26 May 2020

LINEARITY TEST: 50mA

.99990 .99990 .99989 .99989 .99990

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999895  
RSD: 0.0005 % AT 19:56:45 ON 26 May 2020

LINEARITY TEST: 69mA

.99989 .99989 .99989 .99989 .99989

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999890  
RSD: 0.0002 % AT 20:02:28 ON 26 May 2020

LINEARITY TEST: 89mA

.99988 .99988 .99988 .99988 .99987

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999878  
RSD: 0.0004 % AT 20:08:10 ON 26 May 2020

LINEARITY TEST: 109mA

.99987 .99986 .99986 .99985 .99986

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999860  
RSD: 0.0006 % AT 20:13:53 ON 26 May 2020

LINEARITY TEST: 129mA

.99984 .99983 .99983 .99983 .99983

LENGTH OF TIME FOR: CALIBRATION= 25.0 SEC.  
723A AVERAGE CALIBRATION FACTOR: .999830  
RSD: 0.0005 % AT 20:19:36 ON 26 May 2020

LINEARITY TEST: 149mA

.99980            .99980            .99979            .99979            .99978

LENGTH OF TIME FOR: CALIBRATION=            25.0 SEC.

723A AVERAGE CALIBRATION FACTOR:            .999793

RSD: 0.0008 %            AT 20:25:19 ON 26 May 2020

723A MCP DIGITAL INTEGRATOR LINEARITY TEST

LINEARITY TEST: 50mA

.99984            .99985            .99986            .99986            .99987

LENGTH OF TIME FOR: CALIBRATION=            50.0 SEC.

723A AVERAGE CALIBRATION FACTOR:            .999854

RSD: 0.0010 %            AT 20:35:00 ON 26 May 2020

Attachment 4. Instrument Par file that documents the value of the load impedance and calibration resistor.

```

LANL3
2      !Analog Out Device: 1=NI board(Mayak Russia Only); 2=HP34970A
50.05   !Calr Maximum Value
49.95   !Calr Minimum value
50.00125 !Calr (723-A Mockup value) 10mA, JVC 4-30-2020
20.09588 !Lm5:Loadr [ w.K17 energized] JVC 4-30-2020
1.0005   !Mcp Max 1.0005
0.9995   !Mcp Min 0.9995
1      !0=Ask 1=not secret don't ask 2=secret (aliquot weight not entered)
10     !Time_between_i seconds used in Timer_b4_next_i
1      !Print_icc_flag 1=Yes; 0=No
0      !Fe_preox_flag 1=Yes; 0=No
500    !Default Calibration Time in Seconds
0.0002  !Calibration Adjustment Tolerance
0.002   !Calibration out-of-control Tolerance
24     !# Months before LINEARITY warning
12     !# Months before Formal Potential warning
600    !# Months before Noise check warning
1200   !# Months before next automatic Relay Exercise
2.5    !Battery or External Supply baseline 2.5
0      !use DVM Temperature: 0=Not Use; otherwise, use
1.00   !DVM Temperature Coefficient
0      !DVM offset to be set
1.0    !DVM gain to be set
1      !1=Long Log format; otherwise short log format required by MHI
974    !Password, 3-8 Characters, no leading or internal spaces. Needs space at
end before !Comment
0      !Stirrer motor RPM and Temperature connected when = 1
0      !Machine baseline current (uA) to subtract from Cell current

```

Site	SRS	SRS	IAEA	JNC	JNC	MHI#1	MHI#2	MHI#3
Russia	772-1F	723-A	SAL	Sys#1	Sys#2	AH-530	AH-325	AH-505
Mayak								
<hr/>								
!Analog	2=HP	2=HP	2=HP	2=HP	2=HP	2=HP	2=HP	2=HP
1=NI								
!Cal R	100.0066	100.0081	100.0139	100.0116	100.0172	100.0106	100.0115	100.0139
100.????	<---- Values ??							
!LoadR	49.989	49.9663	49.7047	49.9466	49.7723	49.8835	49.8634	49.8133
49.????	<---- Values ??							
!	2001-Jan			2002-Apr	2002-Apr	2002-Apr	2002-Apr	2002-Apr
!MCPMax	1.0002	1.0002	1.0005	1.0005	1.0005	1.0005	1.0005	1.0005
1.0005								
!MCPMin	0.9998	0.9998	0.9995	0.9995	0.9995	0.9995	0.9995	0.9995
0.9995								
!Time	10	10	10	10	10	10	10	10
10								

```
!Secret 1=Not 1=Not 1=Not 1=Not 1=Not 1=Not 1=Not 1=Not
1=Not
!Prt_icc 1=Yes 1=Yes 0=No 1=Yes 1=Yes 0=No 0=No 0=No
0=No
!Fe_preox 0=No 0=No 0=No 0=No 0=No 0=No 0=No 0=No
0=No
!
```

! Password Requirements:

- ! o 3-8 Alpha-Numeric Characters,
- ! o No leading blank spaces
- ! o No internal blank spaces.

! Add a blank space at the end of the password in the parameter file, before providing !Comments...

! This trailing space is NOT part of the password, but is used by the software to define the end.

! Software password default: SRSCOUL