

April 12, 2006

WSRC-MSD-DIS-2004-09663

Mrs Debbie Caver, Technical Information Officer
U. S. Department of Energy - Savannah River Operations Office

Dear Mrs. Caver:

REQUEST FOR APPROVAL TO RELEASE SCIENTIFIC/TECHNICAL INFORMATION

The attached document is submitted for classification and technical approvals for the purpose of external release. Please complete Part II of this letter and return the letter to the undersigned by 5/24/2006. The document has been reviewed for classification and export control by a WSRC Classification staff member and has been determined to be Unclassified.



Kevin Schmidt, Manager
WSRC Technical and Management Information

I. DETAILS OF REQUEST FOR RELEASE

Document Number: WSRC-TR-2006-00060

Author's Name: BOYD HOWARD

Location: 781-A

Phone (803)725-3765

781-A

Document Title: DIGITAL RADIOGRAPHY OF SPECIAL NUCLEAR MATERIAL TEST PACKAGES

Presentation/Publication:

Meeting/Journal: 42nd WANTO Conference

Location: AIKEN SC U.S.A.

Meeting Date: 3/28/2006 to 3/30/2006

Intended Distribution: Unlimited (Release to Public)

II. DOE-SR ACTION

OSTI REPORTABLE

Date Received by TIO _____

- | | |
|--------------------------------------------------------------|--------------------------------------------------------|
| <input type="checkbox"/> Approved for Release | <input type="checkbox"/> Not Approved |
| <input type="checkbox"/> Approved Upon Completion of Changes | <input type="checkbox"/> Revise and Resubmit to DOE-SR |
| <input type="checkbox"/> Approved with Remarks | |

Remarks: _____

Debbie Caver, Technical Information Officer, DOE-SR

Date

THE WSRC TEAM

**Westinghouse Savannah River Company LLC · Bechtel Savannah River, Inc. · BNFL Savannah River Corporation
BWXT Savannah River Company · CH2 Savannah River Company · Polestar Savannah River Company**

Manager's User ID	Y8740	Manager's Full Name (Last, First)	SMAIL, TIMOTHY R
--------------------------	-------	------------------------------------------	------------------

Requested Reviews

WSRC-TR-2006-00060
Classification(if not U)
Unclassified

Submittor Name

BOYD HOWARD

Approval Expiration Date

3/13/2006 Change Expiration Date

Current Reviewer ID and Name

o4304 Shankle, Bob

☒ Classification Review Review all Security

APPROVED by Shankle, Bob on 3/9/2006

☒ OPSEC Review

APPROVED by Shankle, Bob on 3/9/2006

☒ Export Control Review

APPROVED by Shankle, Bob on 3/9/2006

☐ Computer Security Review

REVIEW NOT REQUIRED per Shankle, Bob on 3/9/2006

☐ PAD Review

☐ General Counsel Review

☐ DOE OEA(Giusti) Review

☐ DOE OEA(Ford) Review

Click to View
Document or Folder

Document Link

file:\\msd01\\Mis1\\STIRReview\\WSRC-TR-2

FolderLink

STI Comments

Reviewers Comments

DIGITAL RADIOGRAPHY OF SPECIAL NUCLEAR MATERIAL TEST PACKAGES

by

BOYD HOWARD

Westinghouse Savannah River Company

Savannah River Site

Aiken, South Carolina 29808

Additional Authors:

DOE Contract No. _____

This paper was prepared in connection with work done under the above contract number with the U. S. Department of Energy. By acceptance of this paper, the publisher and/or recipient acknowledges the U. S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering this paper, along with the right to reproduce and to authorize others to reproduce all or part of the copyrighted paper.

Digital Radiography of Special Nuclear Material Test Packages

Boyd Howard, Joel Jones, and Mark Farrar

**FEBRUARY 2006
WSRC-TR-2006-00060**

**Washington Savannah River Company
Savannah River National Laboratory
Aiken, SC 29808**

**DOES NOT CONTAIN
UNCLASSIFIED CONTROLLED
NUCLEAR INFORMATION**

Digital Radiography of Special Nuclear Material Test Packages

Boyd Howard, Joel Jones, and Mark Farrar

**FEBRUARY 2006
WSRC-TR-2006-00060**

**Washington Savannah River Company
Savannah River National Laboratory
Aiken, SC 29808**

Digital Radiography of Special Nuclear Material Test Packages

Objectives

The purpose of this document is to provide a brief introduction to digital radiography (DR), and a description of the DR configuration that was used to radiographically image the Special Nuclear Material (SNM) Test Packages before and after function tests that have been conducted. Also included are (1) Attachment 1, a comprehensive index that describes at which phase of the certification process that digital radiographic images were acquired, (2) digital radiographic images of each of the six packages at various stages of the certification process, and (3) Attachment 2, imaging instructions, that specify the setup procedures and detailed parameters of the DR imaging methodology that were used..

Introduction to Digital Radiography

The diagram to the right shows a typical lens-coupled, CCD-based, area- detector, digital radiography imaging system configuration. The scintillator converts X-ray photons into light. The turning mirror allows the CCD camera to be placed outside of the high-intensity X-ray beam. As the beam of X-rays pass through the test object (in this case a SNM test package), the scintillating screen's rare earth compound interacts with incoming X-ray photons at the atomic level and a visible image appears. The visible image on the scintillation screen is photographically "captured" by an ultra-high resolution digital camera equipped with a scientific grade 2000 by 3000 pixel array imaging chip. The captured image is then stored in a designated work station, after which it is archived on compact discs for evaluation and reporting.




Figure 1: Typical DR Configuration


Digital Radiography of SNM Test Packages

Shown in Figure 2 on the following page is a photo of WSRC's 711-6N test package digital radiography imaging system. The gadolinium oxysulfide (GOS) large area-detector scintillation screen is mounted to the back side of the white panel that is directly behind the mixed waste drum in the photo. The large support panel can accommodate up to twenty (20) 17 inch by 17 inch GOS scintillation screens tiled together than can creat a scintillator 68 inches high and 85 inches wide. The full height of each test package can be imaged in each digital radiographic exposure in this facility.

The platform on which each test package was placed has both vertical and rotational motions. Around the outer circumference of each of the test packages 0, 30, 60, 90, 120, 150, and 180 degrees were marked in a counter clockwise direction, **as viewed from the top of the test package**, thus the


test package rotated through those seven (7) radiographic “views” clockwise, as viewed from the top of the test package.

Lead numbers (0 through 180) were placed on top of the circumferential markings on each test package. Savannah River Packaging Technology had stipulated the zero position.




Attachment 2, SNM Test Package Digital Radiography Instructions, provides specific instructions for positioning the drum on the balsa wood spacer, that is on top of the manipulator table. Those instructions are: “For all SNM TEST PACKAGES and SNM TEST PACKAGE UPPER PORTION images, place the test package vertically on the center of the spacer with the 0° mark facing (near) the scintillation panels.”

Affixed lead number degree markers can be seen in each digital radiographic image, as is illustrated in Figure 3. These images were acquired of a test package before insertion of a containment vessel. In the left test package image the zero (0) degree marker is near the scintillation panels. In the right test package image the thirty (30) degree lead marker is near the scintillation panel. The linear, vertical and horizontal lines in all digital radiograms delineate the outline of the 17 inch by 17 inch gadolinium oxysulfide scintillation screens.



The small, black ovals are intentionally drilled vent holes in the drum.



At the specific request of the Savannah River Packaging Technology group, the upper portion of each of the test packages was digitally radiographed. As can be seen in the two digital radiograms in Figure 4 above, the affixed lead numbers can be seen in each digital radiographic image. These images were acquired of a test package before insertion of a containment vessel. In the left test package image the thirty (30) degree marker is near the scintillation panel. In the right test package image the zero (0) degree lead marker is near the scintillation panel.


There is a vertical, dark line just to the right of the 30 degree lead number in the left digital radiogram. The vertical, dark line was judged to be a separation in the polyurethane foam. That anomaly is also revealed in the end to end digital radiograms in Figure 5.

The linear, vertical and horizontal, white lines in all digital radiograms delineate the outline of each of the 17 inch by 17 inch gadolinium oxysulfide scintillation screens.

In accordance with the attached instructions each of the six (6) drums was digitally radiographed from end to end.


The zero (0) degree reference point is at the top dead center for the end to end exposures.

In Figure 5 are end to end DR images of TP0005. This test package was initially imaged with the



bottom end of the test package near the scintillation panel. After rotating the test package 180 degrees, it was imaged with the top end of the test package near the scintillation panel. The spacing and size of the bolts on the top of the inspection package have not been geometrically enlarged when the top of the test package is nearest the scintillation panel.

The separation in the foam that was noted in the upper portion digital radiograms appear approximately 20 degrees to the left of top dead center. The same separation can be seen approximately 20 degrees to the right of top dead center in the digital radiogram to the right.



A Photometrics digital camera with a scientific grade 2000 by 3000 pixel array, charge-coupled device (CCD) chip was used to acquire all images. Attached to the camera is a 50 mm focal-length Nikon lens. Because the extra large area-detector, digital radiography imaging system is an “open installation,” lead shielding is placed around the camera and its electronics. Optically transparent lead glass, placed between the CCD camera and the turning mirror, blocks scattered X-ray photons that might impinge upon the camera; thus degrading the radiographic image.

A turning mirror is used to focus the CCD camera onto the GOS scintillation screens and to keep the camera out of the direct X-ray beam. Effective resolution of the individual images is approximately 0.030 inches.

The computer controlled acquisition station for the 711-6N DR facility, in which all the Test Packages were radiographically imaged is safely located outside of the imaging enclosure.

Even though the attached instructions directs the Washington Group X-ray specialists to turn off the lights in the 711-6N X-ray imaging room, the imaging room is still not “light tight,” so the entire imaging train, which includes the scintillation screen, the turning mirror, and the lead-shielded CCD camera, is covered with black plastic to exclude unwanted light, as can be seen in Figure x to the right. In the future the imaging train could be enclosed in a more permanent, light-tight structure.



Digital Radiography Configuration and Typical Digital Radiography Imaging Parameters

As can be seen in Figure 1 and Figure y, and specified in the attached Instructions, the distance from the target of the X-ray tube and the scintillation screen is 141 inches. The center of rotation of the rotary stage is as close to the back of the scintillation panel as practicable. The 141 inch X-ray tube focal spot to scintillation screen improves resolution and decreases geometrical enlargement of the object being examined. All radiographic images were acquired at 300 kVp and 5 mA. Each of the seven circumferential exposures of the full test package was 300 seconds. End to end exposure times were 420 seconds.

EVALUATION OF SNM TEST PACKAGES AT EACH STAGE OF IMAGING

GENERAL COMMENTS: It is the writer's professional opinion that the containment vessel in the test packages that were subjected to all of the regulatory-specified tests showed no radiographically detectable signs of degradation.

Note: *Higher density materials such as metal or Fiberfrax appear "white or whitish" in the digital radiograms. Lower density materials such as polyurethane or anomalies in the polyurethane appear to be "shades of grey."*

A complete index of DIGITAL RADIOGRAPHIC IMAGES TAKEN OF SPECIAL NUCLEAR MATERIAL TEST PACKAGES," and each image is in Attachment 1.

TP0002 BEFORE AND AFTER INSERTION OF CONTAINMENT VESSEL (CV)

- There is a vertical separation in the polyurethane foam that appears in 30 and 60 degree images. It does not go the full length of the inspection package.
- Excess Fiberfrax sleeve material is visible beneath the drum liner.

TP0002 AFTER NORMAL CONDITION OF TRANSPORT (NCT) TESTS

- No changes in the containment vessel and/or the polyurethane foam were detected.

TP0002 AFTER THE THIRTY (30) FOOT DROP TEST

- No changes in the containment vessel and/or the polyurethane foam are radiographically detectable when compared with the DR images acquired after the NCT tests.
- There appears to be a difference in the length of dummy load and in the space between the top load distribution fixture and the bottom of the top lid in the before and after images.
- The drum flange has sustained deformation.

TP0002 AFTER THE CRUSH TEST

- The drum suffered some rather severe deformation as a result of the crush test.
- The bottom of the drum is deformed and there is apparent "crumpling" of the drum near the top.
- The configuration of the drum liner lid bottom has been altered; however, the containment vessel itself does not appear to be compromised.
- The polyurethane foam appears to be unchanged.

TP0002 AFTER THE THERMAL TEST

- The configuration of the containment vessel and the drum liner are unchanged and in tact.
- There has been a loss of polyurethane foam; however, polyurethane foam still provides a protective layer over the surface of the drum liner except for a gap between the foam and drum top plate.

TP0003 BEFORE INSERTION OF CONTAINMENT VESSEL (CV)

- The polyurethane foam and drum liner appear to be in acceptable condition; however, there appears to be lower density (voids) in the foam just beneath the drum liner.
- The Fiberfrax sleeve is visible.

TP0003 AFTER THE THIRTY (30) FOOT DROP TEST

- This package was obviously dropped on a bottom edge.
- Quite severe deformation occurred at the point of impact.
- The drum liner in which the CV is placed appears to have “collapsed” slightly; however, the Containment Vessel appears to be unaffected.
- No radiographically detectable damage has occurred in the polyurethane foam.

TP0003 AFTER THE CRUSH TEST

- The polyurethane foam appears to have decreased in density most likely because the bottom of the drum bulged out and possible damage to polyurethane foam.
- The drum suffered some rather severe deformation as a result of the crush test. The bottom of the drum is severely deformed.
-

TP0003 AFTER THE THERMAL TEST

- The configuration of the containment vessel and the drum liner appears to be unchanged and intact.
- Opening in the bottom of drum exposed the polyurethane foam in the bottom section of the drum to the direct flame.
- Direct exposure to the fire has caused a significant loss of polyurethane foam.
- Remaining foam is separated into two lobes around the drum liner.
- Drum liner is only partially covered with undamaged foam.

TP0004 BEFORE INSERTION OF CONTAINMENT VESSEL (CV)

- The internal container appears to be in proper position/orientation...
- There appears to be a separation in the foam, visible only in the 90 degree image.
- The separation appears to be one third (1/3) the height of the drum.
- The separation appears in the end to end images at the 285 degree with the bottom of the test package nearest the scintillator panel.
- The Fiberfrax sleeve is visible beneath the drum liner.

TP0004 AFTER THE THIRTY (30) FOOT DROP TEST

- No changes in the containment vessel and/or the polyurethane foam are radiographically detectable.
- There does not appear to be any obvious physical deformation.

TP0004 AFTER THE CRUSH TEST

- The upper portion of the drum appears to be slightly damaged.
- The configuration of the drum liner lid bottom has been altered.
- The containment vessel itself does not appear to be damaged.
- The polyurethane foam appears to be unchanged.

TP0004 AFTER THE THERMAL TEST

- The configuration of the containment vessel and the drum liner appears to be unchanged and in tact.
- There has been a significant loss of polyurethane foam; however, polyurethane foam still provides a protective layer over most of the surface of the containment vessel.
- The bottom of the drum liner may not be fully covered with polyurethane foam.
- The configuration of the top plug at the top of the inner has not been altered.

TP0005 BEFORE INSERTION OF CONTAINMENT VESSEL (CV)

- The internal container appears to be positioned properly.
- There is a vertical separation in the foam visible in the 30 degree image (1). It may be as much as ½ the entire length of the drum. This separation also appears in the end to end images at approximately 20 degrees counter clockwise from the zero degree mark.
- There are also smaller separations in the polyurethane.

TP0005 AFTER THE THIRTY (30) FOOT DROP TEST

- No changes in the containment vessel and/or the polyurethane foam are radiographically detectable.
- There does not appear to be any obvious physical deformation; however, there is a definite “tilt” to the entire test package.

TP0005 AFTER THE CRUSH TEST

- The containment vessel appears to be unaffected by this test.
- The top and bottom of the drum appear to be quite severely deformed
- The polyurethane foam appears to be unchanged.

TP0005 AFTER THE THERMAL TEST

- The configuration of the containment vessel and the drum liner appears to be unchanged and in tact.
- Direct exposure to the fire has caused a significant loss of polyurethane foam; however, a protective layer of polyurethane foam still covers the upper surface of the containment vessel.
- The configuration of the top plug at the top of the inner has not been altered.

TP0006 BEFORE INSERTION OF CONTAINMENT VESSEL (CV)

- The drum liner appears to be positioned properly.
- There is a vertical separation in the foam visible in several of the images (1). It may be as much as 1/3 the entire length of the drum. This separation also appears in at least one of the end to end images.
- Other separations in the foam have also been imaged such as that just underneath the drum liner.

TP0007 BEFORE INSERTION OF CONTAINMENT VESSEL (CV)

- The internal container appears to be positioned properly.
- There is a vertical separation in the foam visible in the 30 and 60 degree images (1). It may be as much as 1/2 the entire height of the drum. This separation is just barely detectable in one of the end to end images.
- Other separations in the foam have also been imaged such as that just underneath the drum liner.
- The drum liner appears to be positioned properly in the center of the drum.

ATTACHMENT 1
DIGITAL RADIOGRAPHIC IMAGES TAKEN OF SPECIAL NUCLEAR MATERIAL TEST PACKAGES


PAGE	DESCRIPTION OF RADIOGRAPHIC EXAMINATIONS BEFORE AND AFTER EACH CERTIFICATION TEST
15	P0002 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL (CV)
15	TP0002 0, 30, AND 60 DEGREE DR IMAGES UPPER PORTION BEFORE INSERTION OF CV
16	TP0002 90, 120, 150 AND 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
16	TP0002 90, 120, 150 AND 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
17	TP0002 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL
18	TP0002 0, 30, AND 60 DEGREE DR IMAGES AFTER INSERTION OF CV
18	TP0002 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION AFTER INSERTION OF CV
19	TP0002 90, 120, 150, and 180 DEGREE DR IMAGES AFTER INSERTION OF CV
19	TP0002 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION AFTER INSERTION OF CV
20	TP0002 0, 30, and 60 DEGREE DR IMAGES AFTER NCT TESTS
20	TP0002 0, 30, and 60 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER (NORMAL CONDITION DROP (NCT) TESTS
21	TP0002 90, 120, 150, and 180 DEGREE DR IMAGES AFTER NCT TESTS
21	TP0002 90, 120, 150, AND 180 DEGREE DR IMAGES OF THE PACKAGE UPPER PORTION AFTER NCT TESTS
22	TP0002 END TO END DR IMAGES OF THE PACKAGE AFTER NCT TESTS
23	TP0002 0 AND 30 DEGREE DR IMAGES AFTER 30' DROP TEST
24	TP0002 60 AND 90 DEGREE DR IMAGES AFTER 30' DROP TEST
25	TP0002 120, 150 AND 180 DEGREE DR IMAGES AFTER 30' DROP TEST
26	TP0002 END TO END DR IMAGES AFTER 30' DROP TEST
26	TP0002 0 and 30 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST
27	TP0002 60 and 90 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST
27	TP0002 120, 150, and 180 DEGREE DR IMAGES OF THE UPPER PORTION AFTER 30' DROP TEST
53	TP0002 0 AND 30 DEGREE DR IMAGES AFTER CRUSH TEST
53	TP0002 60 AND 90 DEGREE DR IMAGES AFTER CRUSH TESTS
54	TP0002 120, 150, AND 180 DEGREE DR IMAGES AFTER CRUSH TESTS
55	TP0002 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TESTS
55	TP0002 END TO END DR IMAGES AFTER CRUSH TESTS
68	TP0002 0, 30, AND 60 DEGREE DR IMAGES AFTER THERMALTESTS (UPPER IMAGES)
68	TP0002 90, 120, 150 AND 180 DR IMAGES AFTER THERMALTESTS (LOWER IMAGES)
69	TP0002 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES OF THE UPPER PORTION AFTER BURN
69	TP0002 END TO END DR IMAGES AFTER THERMALTEST

PAGE DESCRIPTION OF RADIOGRAPHIC EXAMINATIONS BEFORE AND AFTER EACH CERTIFICATION TEST


28	TP0003 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAIN VESSEL
28	TP0003 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAIN VESSEL
29	TP0003 90, 120, 150 AND 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
29	TP0003 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
30	TP0003 END TO END DR IMAGES BEFORE INSERTION OF CV
31	TP0003 0 AND 30 DEGREE DR IMAGES AFTER 30' DROP TEST
32	TP0003 60 AND 90 DEGREE DR IMAGES AFTER 30' DROP TEST
32	TP0003 120, 150, and 180 DEGREE DR IMAGES OF THE PACKAGE AFTER 30' DROP TEST
34	TP0003 0, 30, AND 60 DEGREE DR IMAGES OF THE UPPER PORTION AFTER 30' DROP TEST
34	TP0003 90, 120, 150, 180 DEGREE DR IMAGES OF THE UPPER PORTION AFTER 30' DROP TEST
35	TP0003 END TO END DR IMAGES AFTER 30' DROP TEST
56	TP0003 0 AND 30 DEGREE DR IMAGES AFTER THE CRUSH TESTS
57	TP0003 60 AND 90 DEGREE DR IMAGES AFTER CRUSH TESTS
58	TP0003 120, 150, AND 180 DEGREE DR IMAGES AFTER THE CRUSH TESTS
59	TP0003 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TEST
59	TP0003 END TO END DR IMAGAGES AFTER CRUSH TESTS
70	TP0003 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES AFTER THERMALTESTS
71	TP0003 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION AFTER THERMALTESTS
71	TP0003 END TO END DR IMAGES AFTER THERMALTEST
36	TP0004 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CV
36	TP0004 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
37	TP0004 90, 120, 150, AND 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
37	TP0004 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
38	TP0004 END TO END DR IMAGES BEFORE INSERTION OF CV
39	TP0004 0, 30, 60, 90, 120, 150, AND 180 DR IMAGES AFTER 30' DROP TEST
40	TP0004 DR IMAGES OF UPPER PORTION AFTER 30' DROP TEST
41	TP0004 END TO END DR IMAGES AFTER 30' DROP TEST
60	TP0004 0 AND 30 DEGREE DR IMAGES AFTER CRUSH TESTS
61	TP0004 60 AND 90 DEGREE DR IMAGES AFTER CRUSH TESTS
62	TP0004 120, 150, AND 180 DEGREE DR IMAGES AFTER CRUSH TESTS
63	TP0004 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TESTS

PAGE DESCRIPTION OF RADIOGRAPHIC EXAMINATIONS BEFORE AND AFTER EACH CERTIFICATION TEST


63	TP0004 END TO END DR IMAGES AFTER CRUSH TESTS
71	TP0004 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES AFTER THERMALTESTS
73	TP0004 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION AFTER THERMALTESTS
73	TP0004 END TO END DR IMAGES AFTER THERMALTEST
41	TP0005 0, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CV
41	TP0005 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
42	TP0005 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
42	TP0005 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
43	TP0005 END TO END DR IMAGES BEFORE INSERTION OF CV
44	TP0005 0, 30, 60, 90, 120, 150, AND 180 DR IMAGES AFTER 30' DROP TEST
45	TP0005 0, 30, 60, 90, 120, 150, AND 180 DR IMAGES OF UPPER PORTION AFTER 30' DROP TEST
45	TP0005 END TO END DR IMAGES AFTER 30' DROP TEST
64	TP0005 0 AND 30 DEGREE DR IMAGES AFTER CRUSH TESTS
65	TP0005 60 AND 90 DEGREE DR IMAGES AFTER CRUSH TESTS
66	TP0005 120, AND 150 DEGREE DR IMAGES AFTER CRUSH TESTS
67	TP0005 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TESTS
67	TP0005 END TO END DR IMAGES AFTER CRUSH TESTS
74	TP0005 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES AFTER THERMALTESTS
75	TP0005 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION AFTER THERMALTESTS
75	TP0005 END TO END DR IMAGES AFTER THERMALTEST
46	TP0006 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CV
46	TP0006 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
47	TP0006 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
47	TP0006 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CV
48	TP0006 END TO END DR IMAGES BEFORE INSERTION OF CV
49	TP0007 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CV
49	TP0007 0, 30, AND 60 DEGREE DR IMAGES OF THE UPPER PORTION BEFORE INSERTION OF CV
50	TP0007 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CV
50	TP0007 90, 120, 150, and 180 DEGREE DR IMAGES OF THE UPPER PORTION BEFORE INSERTION OF CV
51	TP0007 END TO END DR IMAGES BEFORE INSERTION OF CV




TP0002 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAIN VESSEL




TP0002 0, 30, AND 60 DEGREE DR IMAGES UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0002 90, 120, 150 AND 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0002 90, 120, 150 AND 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0002 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0002 0, 30, AND 60 DEGREE DR IMAGES AFTER INSERTION OF CONTAINMENT VESSEL




TP0002 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION AFTER INSERTION OF CONTAINMENT VESSEL




TP0002 90, 120, 150, and 180 DEGREE DR IMAGES AFTER INSERTION OF CONTAINMENT VESSEL




TP0002 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION AFTER INSERTION OF CONTAINMENT VESSEL




TP0002 0, 30, and 60 DEGREE DR IMAGES AFTER NCT TESTS




TP0002 0, 30, and 60 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER NCT TESTS




TP0002 90, 120, 150, and 180 DEGREE DR IMAGES AFTER NCT TESTS



TP0002 90, 120, 150, AND 180 DEGREE DR IMAGES OF THE PACKAGE UPPER PORTION AFTER NCT TESTS



TP0002 END TO END DR IMAGES OF THE PACKAGE AFTER NCT TESTS




0 Degree Image Before 30' Drop Test

0 Degree Image After 30' Drop Test

30 Degree Image After 30' Drop Test

TP0002 0 AND 30 DEGREE DR IMAGES AFTER 30' DROP TEST




60 Degree Image Before 30' Drop Test

60 Degree Image After 30' Drop Test

90 Degree Image After 30' Drop Test

TP0002 60 AND 90 DEGREE DR IMAGES AFTER 30' DROP TEST




120 Degree Image After 30' Drop Test

150 Degree Image After 30' Drop Test

180 Degree Image After 30' Drop

TP0002 120, 150 AND 180 DEGREE DR IMAGES AFTER 30' DROP TEST




End to End DR Image Before Drop Test
Top of Drum Near Scintillator

End to End DR Image After 30' Drop Test
Bottom of Drum Near the Scintillator

End to End DR Image After 30' Drop
Top of Drum Near the Scintillator

TP0002 END TO END DR IMAGES AFTER 30' DROP TEST




0 Degree Image before 30' Drop Test

0 Degree Image After 30' Drop Test

30 Degree Image After 30' Drop Test

TP0002 0 and 30 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST




60 Degree Image After 30' Drop Test

60 Degree Image After 30' Drop Test

90 Degree Image After 30' Drop Test

TP0002 60 and 90 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST




120 Degree Image After 30' Drop Test


150 Degree Image After 30' Drop Test

180 Degree Image After 30' Drop Test


TP0002 120, 150, and 180 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST




TP0003 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0003 0, 30, AND 60 DEGREE DR IMAGES OF THE UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL



TP0003 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0003 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL



End to End DR Image
Bottom of Drum is Nearest the Scintillator

End to End DR Image
Top of Drum is Nearest the Scintillator

TP0003 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




0 Degree Image before 30' Drop Test

0 Degree Image After 30' Drop Test

30 Degree Image After 30' Drop Test

TP0003 0 AND 30 DEGREE DR IMAGES AFTER 30' DROP TEST




60 Degree Image before 30' Drop Test

60 Degree Image After 30' Drop Test

90 Degree Image After 30' Drop Test

TP0003 60 AND 90 DEGREE DR IMAGES AFTER 30' DROP TEST




120 Degree Image After 30' Drop Test

150 Degree Image After 30' Drop Test

180 Degree Image After 30' Drop Test

TP0003 120, 150, and 180 DEGREE DR IMAGES OF THE PACKAGE AFTER 30' DROP TEST




0 Degree Image After 30' Drop Test

30 Degree Image After 30' Drop Test

60 Degree Image After 30' Drop Test

TP0003 0, 30, AND 60 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST




90 Degree Image

120 Degree Image

150 Degree Image

180 Degree Image

TP0003 90, 120, 150, 180 DEGREE DR IMAGES OF THE UPPER PORTION OF THE PACKAGE AFTER 30' DROP TEST




End to End DR Image Before Drop Test
Bottom of Drum Near Scintillator


End to End DR Image After 30' Drop Test
Bottom of Drum Near the Scintillator

End to End DR Image After 30' Drop
Top of Drum Near the Scintillator


TP0003 END TO END DR IMAGES AFTER 30' DROP TEST




TP0004 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0004 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0004 90, 120, 150, AND 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL



TP0004 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




Bottom of the Drum is Nearest the Scintillator




Top of the Drum is Nearest the Scintillator

TP0004 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL


TP0004 DR IMAGES AFTER 30' DROP TEST



UPPER IMAGES: 0, 30, 60, & 90
LOWER IMAGES: 120, 150, and






TP0004 DR IMAGES OF UPPER PORTION AFTER 30' DROP TEST




UPPER IMAGES: 0, 30, 60, AND 90

LOWER IMAGES: 12, 150 AND 180







TP0005 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0005 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0005 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL



TP0005 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




Bottom of the Drum is Nearest the Scintillator



Top of the Drum is Nearest the Scintillator


TP0005 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL

TP0005 DR IMAGES AFTER 30' DROP TEST




**UPPER IMAGES: 0, 30,
60, AND 90**

**LOWER IMAGES: 120,
150, AND 180**



TP0005 DR IMAGES OF UPPER PORTION AFTER 30' DROP TEST



UPPER IMAGES: 0, 30, 60, AND 90

LOWER


IMAGES:

120,


150,


AND

180




**TP0005 END TO END
DR IMAGES**







TP0006 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0006 0, 30, AND 60 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0006 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL



TP0006 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




Bottom of the Drum is Nearest the Scintillator




Top of the Drum is Nearest the Scintillator


TP0006 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0007 0, 30, AND 60 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL




TP0007 0, 30, AND 60 DEGREE DR IMAGES OF THE UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




TP0007 90, 120, 150, and 180 DEGREE DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL



TP0007 90, 120, 150, and 180 DEGREE DR IMAGES OF UPPER PORTION BEFORE INSERTION OF CONTAINMENT VESSEL




Bottom of the Drum is Nearest the Scintillator



Top of the Drum is Nearest the Scintillator


TP0007 END TO END DR IMAGES BEFORE INSERTION OF CONTAINMENT VESSEL



**GPFP0002 0 AND 30
DEGREE DR IMAGES
AFTER CRUSH TEST**

The two upper images were acquired after the crush test.


The lower images were acquired after the 30 foot drop test and are included here for comparison purposes.



**TP0002 60 AND 90
DEGREE DR
IMAGES AFTER
CRUSH TESTS**

The two upper images were acquired after the test that was conducted after the crush tests


The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**



**TP0002 120, 150, AND 180
DEGREE DR IMAGES AFTER
CRUSH TESTS**


The three upper images were acquired after the test that was conducted after the crush tests.

The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**



TP0002 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TESTS


TP0002 END TO END DR IMAGES AFTER CRUSH TESTS



**TP0003 0 AND 30
DEGREE DR IMAGES
AFTER THE CRUSH
TESTS**

The two upper images were
acquired after the crush tests.


The lower images were
acquired after the 30 foot
drop test and **are included
here for comparison
purposes.**




**TP0003 60 AND 90
DEGREE DR IMAGES
AFTER CRUSH TESTS**

The two upper images were acquired after the test that was conducted after the crush tests.


The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**






TP0003 120, 150, AND 180 DEGREE DR IMAGES AFTER THE CRUSH TESTS


The three upper images were acquired after the test that was conducted after the crush tests. The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**



GPFP0003 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TEST

**TP0003 END TO END DR
IMAGAGES AFTER
CRUSH TESTS**






**TP0004 0 AND 30
DEGREE DR
IMAGES AFTER
CRUSH TESTS**

The two upper images were acquired after the crush tests.


The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**



**TP0004 60 AND 90
DEGREE DR IMAGES
AFTER CRUSH TESTS**

The two upper images were
acquired after the crush tests.


The lower images were
acquired after the 30 foot drop
test and **are included here for
comparison purposes.**




**TP0004 120, 150, AND 180
DEGREE DR IMAGES AFTER
CRUSH TESTS**

The three upper images were
acquired after the crush tests.


The lower images were acquired after
the 30 foot drop test and **are
included here for comparison
purposes.**



TP0004 0, 30, 60, 90, 120, 150, AND 180 DEGREE UPPER PORTION DR IMAGES AFTER CRUSH TESTS




**TP0004 END TO
END DR IMAGES
AFTER CRUSH
TESTS**



**TP00005 0 AND
30 DEGREE
DR IMAGES
AFTER
CRUSH TESTS**

The two upper images were acquired after the crush tests.


The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**



**TP0005 60 AND
90 DEGREE
DR IMAGES
AFTER
CRUSH TESTS**

The two upper images were acquired after the crush tests.


The lower images were acquired after the 30 foot drop test and **are included here for comparison purposes.**




**TP00005 120
AND 150
DEGREE DR
IMAGES
AFTER CRUSH
TESTS**

The two upper
images were
acquired after the
crush tests. No
180 DR.


The lower images,
TP0005, 120, 150,
and 180 were
acquired after the
30 foot drop test
and **are included
here for
comparison
purposes.**




TP0005: 0, 30, 60, 90, 120, 150, AND 180° UPPER PORTION IMAGES AFTER CRUSH TESTS






TP0005 END TO END DR IMAGES AFTER CRUSH TESTS




TP0002 0, 30, AND 60 DEGREE DR IMAGES AFTER THERMAL TESTS (UPPER IMAGES)




TP0002 90, 120, 150 AND 180 DR IMAGES AFTER THERMAL TESTS (LOWER IMAGES)




**TP0002 0, 30, 60, 90, 120, 150, & 180° DR IMAGES OF THE UPPER PORTION AFTER THERMAL
TP0002 END TO END DR IMAGES AFTER THERMAL TESTS**




TP0003 0, 30 AND 60 DEGREE DR IMAGES AFTER THERMALTESTS (UPPER IMAGES)




TP0003 90, 120, 150 AND 180° DR IMAGES AFTER THERMALTESTS (LOWER IMAGES)




TP0003 0, 30, 60, 90, 120, 150, AND 180° DR IMAGES OF UPPER PORTION AFTER THERMAL TESTS




TP0003 END TO END DR IMAGES AFTER THERMAL TEST




TP0004 0, 30 AND 60 DEGREE DR IMAGES AFTER THERMAL TESTS (UPPER IMAGES)




TP0004 90, 120, 150, AND 180 DEGREE DR IMAGES AFTER THERMAL TESTS (LOWER IMAGES)




TP0004 0, 30, 60, 90, 120, 150, AND 180° DR IMAGES OF UPPER PORTION AFTER THERMAL TEST





TP0004 END TO END DR IMAGES AFTER THERMAL TEST



TP0005 0, 30 AND 60 DEGREE DR IMAGES AFTER THERMAL TESTS (UPPER IMAGES)




TP0005 90, 120, 150 AND 180 DR IMAGES AFTER THERMAL TESTS (LOWER IMAGES)



TP0005 0, 30, 60, 90, 120, 150, AND 180 DEGREE DR IMAGES OF UPPER PORTION AFTER THERMALTEST

**TP0005 END TO END DR
IMAGES AFTER
THERMALTEST**

(ONLY THE IMAGE IN
WHICH THE TOP OF THE
PACKAGE IS NEAREST THE
SCINTILLATOR WAS
ARCHIVED.)





SRNL

SAVANNAH RIVER NATIONAL LABORATORY

ATTACHMENT 2

Engineered Equipment and Systems

Imaging Instructions

SNM Test Package Digital Radiography



Imaging Instructions

Contacts:

Joel D. Jones
WSRC/SRNL/EES
Bldg 781-A
Aiken, SC 29808
803-725-3672 /P19449
joel.jones@srnl.doe.gov

Mark Farrar
WSRC/SRNL/ EES
Bldg 730-A
Aiken, SC 29808
803-725-1786/P13128
mark.farrar@srnl.doe.gov

Boyd Howard
WSRC/SRNL/EES
Bldg 781-A
Aiken, SC 29808
803-7253765/P18782
boyd.howard@srnl.doe.gov

SNM TEST PACKAGE Digital Radiography

Engineered Equipment and Systems

About Us...

The Engineered Equipment and Systems Department (EES) serves the Savannah River Site (SRS) operations, other Department of Energy (DOE) Sites, and the private sector primarily where **off-the-shelf design cannot fill a need**. We are an **applied engineering house**, employing electrical, mechanical and computer engineering techniques to solve unique customer needs where measurement, fixturing, or inspection is required but limitations in either field conditions or available technology requires a novel solution. IES has experience in modifying or refining commercial equipment, or designing new devices to meet unique design requirements.

EES is a branch of the Savannah River National Laboratory (SRNL) which is operated for the U.S. Department of Energy by Westinghouse Savannah River Company. Our personnel include Mechanical / Electrical engineers, NDE engineers, designers, drafters, Engineering Assistants, certified leak test personnel and more.

MISSION

Our Mission is to provide technical and engineering development services to SRS operations, other DOE sites, and the private sector in support of the DOE mission.

VISION

To be recognized as the most effective leader in the development and application of technology for the DOE.

Contact Us At

<http://shrine.srs.gov/html/srnl/dept/eesd/main.html>

EES Department Manager: (803) 725-2331

Revision History

Revision	Date	Author	Summary
0	8/4/2005	J. D. Jones	Original issue

About This Document

This document provides instructions for performing Digital Radiography on SNM TEST PACKAGES in N-Area. It provides instructions to operate the computer and digital imaging software. It also provides instructions for setting up the test packages for imaging.

This document **does not** provide instructions for operating the X-Ray machine.

This document **is not** intended to be a step by step procedure. It is a **reference guide** to be used by personnel familiar with the operation of the system.

Conventions Used in This Manual

The following conventions are used in this manual:

>>

The >> symbol leads you through nested menu items and dialog box options to a final action. The sequence **SHOW>>Temperature Graph** directs you to pull down the **SHOW** menu and select the Temperature Graph option.

G

This icon denotes a tip, which alerts you to advisory information.

!

This icon denotes a note, which alerts you to important information.

bold

Bold text denotes items that you must select or click on in the software, such as menu items and dialog box options. **Bold text** also denotes parameter names.

italic

Italic text denotes variables, emphasis, a cross reference, or an introduction to a key concept. This font also denotes text that is a placeholder for a word or value that you must supply.

Microsoft
Sans Serif

Text in this font denotes text or characters that you should enter from the keyboard, sections of code, programming examples, and syntax examples. This font is also used for the proper names of disk drives, paths, directories, programs, subprograms, subroutines, device names, functions, operations, variables, filenames and extensions, and code excerpts.

Contents

1	X-RAY ROOM	1
1.1	TURN POWER ON	1
1.2	POSITION/PREPARE THE TURNTABLE	1
1.3	POSITION TEST PACKAGE	1
1.4	POSITION THE X-RAY HEAD	2
1.5	TURN OUT LIGHTS	2
2	IMAGING SOFTWARE	3
2.1	TURN POWER ON	3
2.2	START THE IMAGING SOFTWARE	3
2.3	SESSION SELECTION	4
2.4	SESSION CONTROL	5
2.5	STAGE CONTROL	8
2.6	SERIES IMAGE	9
2.6.1	Setup Session Series	9
2.6.2	Take Session Series	11
2.7	SINGLE IMAGE	12
2.7.1	Preview Image	12
2.7.2	Final Image	13
3	PARAMETER SUMMARY	14
3.1	X-RAY PARAMETERS	14
3.2	GEOMETRIC PARAMETERS	14
3.3	FILE NAMING	15

16 X-Ray Room

This section provides instructions for setting-up a SNM TEST PACKAGE in the 711-6N, N-Area X-ray Room

Turn Power On

Turn Power On to the equipment as follows:

- ✦ Turn outlet strip adjacent to the turntable on. Indicator lights will illuminate if power is available to the strip. This strip provides power to the digital camera, the turntable, and the turntable servo drive control box.
- ✦ Plug power into the back of the video camera adjacent to the servo drive control box.

Position/Prepare the Turntable

Position and prepare the Turntable Assembly as follows:

- ✦ Place the turntable close (within 4") to the scintillation screen without touching it. It must be able to rotate freely. Center the turntable with the center of the Scintillation Screen.
- ✦ Position the top edge of the turntable approximately 20.5" from the floor.
- ✦ Place the round white plastic spacer in the center of the turntable. The spacer is 3.5" high.

Position Test Package

Position the Test Package on the Turntable as follows:

- ✦ For all SNM TEST PACKAGES and SNM TEST PACKAGE upper portion images, place the test package vertically on the center of the spacer with the 0° mark facing the scintillation screen. The 0-180° Axis should be approximately perpendicular to the screen.
- ✦ For the SNM TEST PACKAGE END to END TOP image, position the test package horizontally on the spacer with the top of the test package closest to the scintillation screen and the bottom towards the X-Ray source. The 0° mark should be facing upwards opposite the spacer.

- For the SNM TEST PACKAGE END-TO-END BOTTOM image, position the test package horizontally on the spacer with the bottom of the test package closest to the scintillation screen and the top towards the X-Ray source. The 0° mark should be facing upwards opposite the spacer.
- Place the number belt on the test package ~1.5” above the first chime (support band) just below the test package serial number. The 0 number on the belt should be aligned with the 0° marking on the test package. The belt runs from 0° to 180° counter clockwise.

Position the X-Ray Head

Position the X-Ray Head as follows:

- Move the X-Ray head horizontally as far back as it will travel away from the turn-table. The head should be approximately 141” from the front of the scintillator screen.
- Position the X-Ray head with the focal spot approximately perpendicular to the scintillator screen.
- For SNM TEST PACKAGE ALL position the center of the X-Ray head focal spot approximately 43” from the floor.
- For SNM TEST PACKAGE upper portion images, position the center of the X-Ray head focal spot approximately 54” from the floor.
- For SNM TEST PACKAGE END-to-END images, position the center of the X-Ray head focal spot approximately 34.5” from the floor (as far down as it will go).

Turn Out Lights

Prior to leaving the X-Ray Room to take images:

- Ensure that the black plastic tent is in place around the Digital Camera, turning mirror and Scintillation screen.
- Turn out all lights.

17 Imaging Software

This section provides instructions for preparing the control room for SNM TEST PACKAGE Imaging and operating the imaging software. This section does not provide exhaustive instructions for the software. It provides the most prominent features used for imaging the test packages and should provide a useful reference for personnel familiar with the systems operation.

Turn Power On

Turn Power On to the equipment as follows:

- Authorized and trained personnel will operate the X-Ray equipment per established procedures not in the scope of this document.
- Turn the video monitor power on.
- Turn the computer power on by pressing the round power button on the front of the computer in the lower right-hand corner.
- When the Log-In prompt appears, log in as user – Administrator without a password. Press the [Enter] key.

Start the Imaging Software

Once the Windows desktop appears and the computer has completed start-up, start the Digital Radiography Imaging Application as follows:

- Double click the Start DR Imaging Icon on the desktop. This launches the LabVIEW application found at C:\SRS DR Software\SRS\start session.vi.
- When the application starts the session selection window will be displayed.

Session Selection

When the application is started the initial application window is displayed. It will appear similar to the following:

The screenshot shows the 'Digital Radiography Imaging' application window. It features a 'Sessions' list on the left with three entries: 'GPFP_Drum_Top', 'GPFP_Drum_End', and 'GPFP_Drum_All'. A red text instruction says 'Double Click on Session Name to Enter'. The 'Session' section at the top center displays 'GPFP_Drum_All', the date 'Wednesday, August 03, 2005', and the time '3:07 PM'. Below this is the 'Acquisition Parameters' section, which includes 'Exposure Time (Sec)' set to '300.00', 'Average' set to '1', 'Image Binning' set to '2', and 'Preview Binning' set to '4'. There are also 'Imager Type' (set to 'CCD') and 'Display Orientation' (set to '90') options. A 'Start New Session' button and an 'Exit' button are at the bottom center. On the right side, there are three yellow panels: 'Job Information' with fields for 'Job Number' (GPFP_A200508), 'Job Title' (GPFP DRUM ALL), and 'Procedure Number'; 'Position Information' with fields for 'Xray Head Distance', 'Pixel Size' (.015), and 'Rotation Center Distance'; and 'Xray Information' with fields for 'Potential' (300 kV), 'Current' (5 mA), 'Spot Size' (4.5 mm), 'Xray Filter Description' (Cu + Brass), and 'Scintillator Description'. A copyright notice 'Copyright SRTC/W5RC Ver 1.0 12/2001' is at the bottom right.

- ➔ A list of the previously created imaging sessions will appear in the **Sessions** window. select the appropriate session as follows:
 - For SNM TEST PACKAGE “ALL” images, Double click the SNM TEST PACKAGE_TEST PACKAGE_ALL session.
 - For SNM TEST PACKAGE TOP images, Double click the SNM TEST PACKAGE_TEST PACKAGE_TOP session.
 - For SNM TEST PACKAGE END images, Double click the SNM TEST PACKAGE__END session.
- ➔ After you select the session control window is displayed.

Session Control

When the Session Control window is displayed, the appropriate parameter values are automatically filled in for the selected session.

- ➔ For SNM TEST PACKAGE ALL images, verify that the Session Control Window values are as follows:

The screenshot shows the 'DR Session Control' window with the following sections and values:

- Session:** GPFP_Drum_All
- Imager Attributes:**
 - Name: Camera1
 - Column (Serial) Size: 3072
 - Row (Parallel) Size: 2048
 - Bit Depth: 4096
- Acquisition Parameters:**
 - Exposure Time (Sec): 300.00
 - Average: 1
 - Gain: 1
 - Image Binning: 2
 - Preview Binning: 4
- Job Information:**
 - Job Number: GPFP_A200508
 - Job Title: GPFP DRUM ALL
 - Procedure Number: (empty)
- Position Information:**
 - Xray Head Distance: (empty)
 - Pixel Size: .015
 - Rotation Center Distance: (empty)
- Xray Information:**
 - Potential: 300 kV
 - Current: 5 mA
 - Spot Size: 4.5 mm
 - Xray Filter Description: Cu + Brass
 - Scintillator Description: (empty)
- Control Buttons:**
 - Got Light: Take Light
 - Got Dark: Take Dark
 - Display Stage Control
 - Done
 - Set Active Region
 - Image
 - Series Image
 - Display Orientation: 90
- Crop:**
 - Top: 232
 - Left: 140
 - Right: 72
 - Bottom: 120

- ➔ For SNM TEST PACKAGE TOP images, verify that the Session Control Window values are as follows:

The screenshot shows the 'DR Session Control' window with the following settings:

- Session:** GPFP_Drum_Top
- Imager Attributes:**
 - Name: Camera1
 - Column (Serial) Size: 3072
 - Row (Parallel) Size: 2048
 - Bit Depth: 4096
- Acquisition Parameters:**
 - Exposure Time (Sec): 360.00
 - Average: 1
 - Gain: 1
 - Image Binning: 2
 - Preview Binning: 4
 - Crop:** Top: 352, Left: 1860, Right: 264, Bottom: 260
- Job Information:**
 - Job Number: GPFPDRUM200508
 - Job Title: GPFP DRUMS Top Section Images
 - Procedure Number: (empty)
- Position Information:**
 - Xray Head Distance: (empty)
 - Pixel Size: .015
 - Rotation Center Distance: (empty)
- Xray Information:**
 - Potential: 300 kV
 - Current: 5 mA
 - Spot Size: 4.5 mm
 - Xray Filter Description: .010 Cu +hin Cu/Brass
 - Scintillator Description: (empty)
- Controls:**
 - Got Light: (button)
 - Got Dark: (button)
 - Take Light: (button)
 - Take Dark: (button)
 - Display Stage Control: (button)
 - Done: (button)
 - Set Active Region: (button)
 - Image: (button)
 - Series Image: (button)
 - Display Orientation: 90


- For SNM TEST PACKAGE END-to-END images, verify that the Session Control Window values are as follows:

The screenshot shows the 'DR Session Control' window with the following sections and values:

- Session:**
 - GPFP_Drum_End
- Imager Attributes:**
 - Name: Camera1
 - Column (Serial) Size: 3072
 - Row (Parallel) Size: 2048
 - Bit Depth: 4096
- Acquisition Parameters:**
 - Exposure Time (Sec): 420.00
 - Average: 1
 - Gain: 1
 - Image Binning: 2
 - Preview Binning: 4
- Job Information:**
 - Job Number: GPFPE200508
 - Job Title: GPFP Drum End Shots
 - Procedure Number: (empty)
- Position Information:**
 - Xray Head Distance: (empty)
 - Pixel Size: .015
 - Rotation Center Distance: (empty)
- Xray Information:**
 - Potential: 300 kV
 - Current: 5 mA
 - Spot Size: 4.5 mm
 - Xray Filter Description: .010 Cu + Brass
 - Scintillator Description: (empty)
- Control Buttons and Indicators:**
 - Got Light: (green dot) Take Light
 - Got Dark: (green dot) Take Dark
 - Display Stage Control
 - Done
 - Set Active Region
 - Image
 - Series Image
 - Display Orientation: 90
- Crop Window:**
 - Top: 48
 - Left: 100
 - Right: 1092
 - Bottom: 8

Stage Control

When the Session Control window is displayed, the Stage Control (Move Table. I) window used to control the turntable is also displayed. If it is not displayed, then select the [Display Stage Control] button to display it.



- Prior to performing a series of images, it usually best to select the [Reset] button to reset the turntable's servo drive controller. It takes approximately 10 seconds for the reset to complete.
- After the turntable drive is reset, Enter the current position of the turntable relative to the test packages position to the scintillator screen in the Enter Degrees: entry box. Usually this value will be zero prior to performing a series of images.
- Select the [SetPos] button. The current position will be set to the position in the Enter Degrees: entry box.

Series Image

In the Session Control window select the [Series Image] button to begin a series of images on the Entire Test package or the Test package Top. The Setup Session Series Window is displayed.

Setup Session Series

- For SNM TEST PACKAGE ALL images, the Setup Session Series Window will appear as follows:


The screenshot shows the 'Setup Session Series.vi' window with the following settings:

- Session:** GPFP_Drum_All
- Image Name:** GPFP0006_AEB
- Imager Attributes:**
 - Name: Camera1
 - Column (Serial) Size: 3072
 - Row (Parallel) Size: 2048
 - Bit Depth: 4096
- Image Region:**
 - Top: 70
 - Left: 116
 - High: 1499
 - Wide: 963

Buttons: Set Region
- Start Position:**
 - TurnTable: 0.0
 - X Axis: 0.00
 - Y Axis: 0.00
 - Z Axis: 0.00


Buttons: Display Stage Control, Go, Cancel
- Motion:**
 - Axis: T
 - Speed: 1.00
 - Range: 210.00
 - Images: 7
 - Step Size: 30.00
- Acquisition Parameters:**
 - Exposure Time (Sec): 300.00
 - Average: 1
 - Gain: 1
 - Image Binning: 2
 - Preview Binning: 4

- For SNM TEST PACKAGE Top images, the Setup Session Series Window will appear as follows:




- In the Setup Session Series window
 - Enter the Image Name as described in the Parameter Summary Section of this guide.
 - Verify that the other values are as shown above.
 - Press the [Go] button. The Take Session Series window will be displayed.

Take Session Series



- ➔ After the X-Ray source is ready, Press the [Resume] button to begin taking the series of images.
- ➔ After each image is taken, the image will be displayed on the screen. Examine the image to verify that the position is correct and that it is changing after each shot. You can use the display loop.vi window controls to adjust the image for better viewing.




Single Image

In the Session Control window select the [Image] button to take a single image.


Preview Image

After selecting the [Image] button. The application takes a preview image.

- The Exposure Meter window is displayed while the preview is taken. The Exposure Meter is window is shown anytime an image is being acquired.



- After the image is acquired, the image is displayed along with the Display Image Controls window.




- You can use the controls to improve the displayed image.
- You can write a bitmap of the image.

- Press [Done] if you are ready to take the final image or Press [Cancel] if you need to make adjustments before taking the final image.


Final Image

If you select the [Done] button after the preview image is taken, the application takes the final image.

- ➔ The Exposure Meter window is displayed while the image is taken.



- ➔ After the image is acquired, the image is displayed along with the Display Image Controls window.



- You can use the controls to improve the displayed image.
- You can write a bitmap of the image.
- **To save the raw image data**, Enter the Image Name prior to pressing [Done]. *If you do not enter a name or if you press [Cancel] the raw image data will not be saved.*

18 Parameter Summary

This section provides a summary of various parameters for each type of SNM TEST PACKAGE Test package Shot.

X-Ray Parameters

The following table summarizes the X-Ray parameters used for the SNM TEST PACKAGE Shots:

Parameter	Test package All	Test package Top	Test package End
Source Intensity	300 kV / 5 mA	300 kV / 5 mA	300 kV / 5 mA
Exposure Time	300 Sec	360 Sec	420 Sec
Num Shots	7 – 0-180° @ 30° increments	7 – 0-180° @ 30° increments	2 – 1 each end

Geometric Parameters

The following table summarizes the X-Ray parameters used for the SNM TEST PACKAGE Shots (Dimensions shown are approximate):

Parameter	Test package All	Test package Top	Test package End
TurnTable Ht (Floor to Top)	20.5''	20.5''	20.5''
X-Ray Src Ht (Floor to Focal Spot)	43''	54''	34.5'' (All the way down)
X-Ray Src Focal Spot to North Wall	47.5''	47.5''	47.5''
X-Ray Src to Screen	141''	141''	141''
X-Ray Src to Test package Center	123''	123''	123''

File Naming

The raw image files should be named as follows:

- For SNM TEST PACKAGE ALL images, The name prefix should be SNM TEST PACKAGE####_Asuffix_seqNum where #### is the serial number of the test package. The test package suffix and seqNum will be described below. For SNM TEST PACKAGE TOP images, The name prefix should be SNM TEST PACKAGE####_Tsufffix where #### is the serial number of the test package. The test package suffix will be described below.
- For SNM TEST PACKAGE TOP END images, The name prefix should be SNM TEST PACKAGE####_THsufffix where #### is the serial number of the test package. The test package suffix will be described below.
- For SNM TEST PACKAGE BOTTOM END images, The name prefix should be SNM TEST PACKAGE####_BHsufffix where #### is the serial number of the test package. The test package suffix will be described below.
- The name suffixes should be as follows:
 - EB - for Empty Baseline
 - CVB - for Containment Vessel inserted Baseline.
 - TA - First Test package Test - Sandia Vibration Analysis
 - TB - Define
 - TC - Define
 - TD - Define
 - TE - Define
 - TF -Define
 - TG -Define
- The seqNum is the sequence number for a series of images. The _# is automatically appended to the Series Image Name for each image in the series. # is as follows:

> 0 – 0° shot	> 4 – 120° shot
> 1 – 30° shot	> 5 – 150° shot
> 2 – 60° shot	> 6 – 180° shot
> 3 – 90° shot	

Disclaimer

This report was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor any agency thereof, nor any of their employees, makes any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy of any information, apparatus, product or process disclosed, or represents that its use would not infringe privately owned rights. References herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring the United States Government or any agency thereof. The views and opinions of the author expressed herein do not necessarily state or reflect those of the United States Government or any agency thereof.